



# Fujian Jixiang Medical Group

福建集翔医疗科技有限公司  
Fujian Jixiang Medical Technology Co., Ltd

1

# Jixiang

## Medical Group

地理位置 Location

# 长汀县（汀州府城）

简称“汀”，隶属福建省，是中央苏区汀州市所在地

[1]、红旗跃过汀江的地方

[2]、位于福建省西部，武夷山脉南麓，南邻广东、西接江西；自古为闽、粤、赣三省边陲要冲，被誉为“福建省西大门”。

[3]、长汀历称汀州，汉代置县，自唐开元二十四年置汀州始，成为福建五大州、七闽地、八闽府之一。自盛唐以来的一千多年里，长汀一直是州、郡、路、府治所在地，是“客家文化（闽西）生态保护实验区”的科教文化中心。

[4]、长汀亦称汀州府，是客家人聚居的第一座府治城市，被誉为“世界客家首府”。长汀还是国务院批复的第三批国家历史文化名城（1994年）、中国烹饪协会首个颁布的中国客家菜之乡（2004年），享有“福建美食名城”、“红军的故乡”、“红色小上海”的称号。

[5]、长汀融人文景观与自然景观于一体，被“中国的十大国际友人”路易·艾黎誉为“中国最美丽的山城”。

[6]、2012年获“中国十大最具人文底蕴古城古镇”称号。长汀成为全国第一批“绿水青山就是金山银山”实践创新基地、首批国家生态文明建设示范市（县），成为福建省唯一同时获得这两个称号（授牌）的县（市）。2018年3月成为第一批通过全国水生态文明建设试点验收城市。

[7]、2019年3月，被列为第一批革命文物保护利用片区分县名单。

[8]、2019年度福建省县域经济发展“十佳”县（市）。

[9]、2020年5月，被列入国家级县城新型城镇化建设示范名单。

2019年，全县地区生产总值达（以四经普为基数核算）289.75亿元，按可比价格计算，增长7.4%

# Changting County (Tingzhou Fucheng)

Referred to as "Ting", belongs to Fujian Province, is the seat of The Central Suzhou District Tingzhou City

[1] Where the red flag jumps over the Ting River.

[2] It is located in the west of Fujian Province, at the southern foot of Wuyi Mountain, adjacent to Guangdong in the south and Jiangxi in the west. Since ancient times, it has been the border crossing of Fujian, Guangdong and Jiangxi province, and is known as the "western gate of Fujian province".

[3] In the Changting calendar, Tingzhou was established as a county in the Han Dynasty. Since it was established in The 24th year of the Kaiyuan of the Tang Dynasty, It has become one of the five provinces of Fujian, the seven provinces of Fujian and the eight Provinces of Fujian. For more than one thousand years since the tang Dynasty, Changting has been the seat of prefecture, county, road and government, and the scientific, educational and cultural center of "Hakka Culture (Western Fujian) Ecological Protection Experimental Area".

[4] Changting, also known as Tingzhou Capital, is the first prefecture city inhabited by Hakka people and is honored as the "Hakka capital of the world". Changting is also the third batch of national historical and cultural cities approved by the State Council (1994), the first Chinese Hakka cuisine issued by the Chinese Culinary Association (2004), and enjoys the title of "Fujian cuisine city", "hometown of the Red Army", "Red Little Shanghai".

[5] Changting integrates cultural landscape with natural landscape, and is praised as "The most beautiful mountain city in China" by "China's Top ten International Friends" Louis Reille.

[6] In 2012, it won the title of "Top Ten Ancient Towns with The Most Cultural Deposits in China". Changting became the first batch of practice and innovation bases of "clear water and green mountains are gold and silver mountains" and the first batch of demonstration cities (counties) of national ecological civilization construction, and became the only county (city) in Fujian Province to win both of these titles (awarding). In March 2018, it became the first batch of cities to pass the national pilot inspection of water ecological civilization construction.

[7] In March 2019, it was listed as one of the first counties for the protection and utilization of revolutionary cultural relics.

[8]. "Top ten" counties (cities) in County economic Development of Fujian Province in 2019.[9] In May 2020, it was included in the demonstration list of new-type urbanization construction of state-level counties. In 2019, the county's gross domestic product (GDP) reached 28.975 billion yuan (calculated from the base of the four jingpu), an increase of 7.4% at comparable prices

2

# **Jixiang** Medical Group

选择长汀 Choose the Changding

# 医疗无界限 -

## 中国首位开国医学专业中將 傅连璋

1894年出生于长汀县河田镇伯公岭村,1925年被推举为长汀县福音医院院长。红军起义初期无偿收治红军伤员,红军长征期间数次为军中发生的各种传染病提出解决办法和医治方案,并主张为红军官兵接种疫苗,避免病情的扩散。此后更担任毛主席保健医生,多次有效维护了毛主席身体健康,免于传染病与疾病的侵扰。新中国成立后历任中央卫生部副部长,中华医学会会长等重要职务,为中国传染病防治体系奠定了不可磨灭功绩。

2020年新冠疫情肆虐全球,长汀县政府迅速筹建『防疫物资生产基地』,提供给防疫物资生产企业多项优惠政策和扶持,本着『人溺己溺』的慈悲胸怀积极招揽各地优质生产企业,为提供给全球各地迫切需要的防疫物资做出充分努力,这无疑是将长汀『**医疗无界限**』的精神升华的体现。福建集翔医疗集团于是选择落地长汀,为这场人类文明史上有记载以来最大的传染病疫情投入防疫物的生产。



傅连璋院长

# Medical treatment knows No boundaries –

China's first pioneer of medical science, Lt. Gen. - Lian Zhang FU

He was born in Bogongling Village, Hetian Town, Changting County in 1894, and was elected President of Changting County Evangelion Hospital in 1925. In the early days of the Red Army uprising, the wounded were admitted and treated free of charge. During the Long March, the Red Army provided solutions and medical treatment for various infectious diseases in the army on several occasions, and advocated vaccinating the Red Army officers and soldiers to prevent the spread of diseases. Since then, he also served as chairman MAO's health care doctor, and repeatedly effectively maintained Chairman MAO's health, free from infectious diseases and diseases. Since the founding of the People's Republic of China, he has served successively as vice Minister of the Central Health Ministry and President of the Chinese Medical Association, which have made ineffable contributions to China's infectious disease prevention and control system. Around the world in 2020 outbreak of the new champions league and changting county government's rapid preparation "epidemic prevention materials production base", provide epidemic prevention material production enterprises a number of preferential policies and support, in line with the "people drowned oneself drowned" compassion around the mind actively attract high-quality production enterprises, to provide the world desperately needs full efforts of epidemic prevention materials, this is undoubtedly the changting "health without boundaries," the spirit of sublimation. Fujian Jixiang Medical Group then chose to land in Changting, for the history of human civilization since the largest recorded epidemic of infectious diseases into the production of epidemic prevention products.

# 3

## Company Profile

企业概况 Enterprise brief introduction

# Fujian Jixiang Medical Technology Co., Ltd

福建集翔医疗科技有限公司

Fujian Jixiang Medical Technology Co., Ltd. (hereinafter referred to as the Company) is a large-scale medical devices enterprise with the funding of China National Pharmaceutical Group Co., Ltd. (Sinopharm) and Xiamen Hansen Group Co., Ltd., which is introduced by the People's Government of Changting County, Fujian Province

**福建集翔医疗科技有限公司(以下简称公司)是一家由福建省长汀县人民政府招商引资，由国药集团和厦门瀚森集团投资建设的大型医疗器械类企业，**

Founded in August, 2020, the Company is located in No. 29 Tingzhou Avenue, Cewu Town, Changting County, Fujian Province, boasting a total area of 170,000 square meters and a plant of 50,000 square meters. At present, its main products have been medical and household gloves made of PVC, latex, nitrile, etc.. Relying on the registered capital of 100 million RMB and the total investment of 1,050 million RMB, the Company plans to build 39 large-scale automatic production lines with a length of 180 meters. By then, with the joint efforts of more than 100 technicians and 800 employees, the Company is able to reach an annual output of 7 billion gloves, paying a tax of at least 100 million RMB each year. The scale of production and sales ranks the first in Fujian Province.

公司成立于2020年8月，坐落于福建长汀县策武镇汀州大道29号，占地面积为17万平方米，厂房面积50000平方米，目前主要产品为PVC、乳胶、丁腈等医用手套民用手套。公司注册资本为1亿元人民币，第一阶段总投资为10.5亿元人民币，公司共计划投建200条丁腈及乳胶生产，300条混丁合成及PVC生产线。第一批次将建成39条180米长的大型全自动生产线，60条全新混丁合成及PVC产线，届时将年产450亿只手套，各项专业技术人员200余人，员工2800余人，年纳税额将超过2亿元人民币以上。在同行业中产、销规模全省第一。

# Fujian Jixiang Medical Technology Co., Ltd

福建集翔医疗科技有限公司

In the spirit of unity, innovation, excellence, cooperation and leadership, the Company has been adhering to the humanized management, upholding the business philosophy of science and technology constitute a primary productive force, and regarding R&D and innovation as the cornerstone to improve the competitiveness. So far, all production lines are independently developed by the Company, which reflects its scientific and technological strength. Sticking to the marketing strategy of dwelling on domestic markets and exploiting overseas markets, the Company has set up a long-term strategic goal of promoting international cooperation. Its products now find a good sale in Southeast Asia, Africa, Europe and other regions.

**团结、创新、卓越、共赢、领航是公司的宗旨和精神，公司始终秉持人性化的管理理念，以科技是第一生产力作为企业发展的主调，永远把研发和创新当着企业提高竞争力的基石。目前所有生产线，皆为公司自主研发，拥有较高的技术含量。公司以“站稳国内，开拓国际市场”积极推进国际化作为企业长远的战略发展目标。并逐步实现产品远销东南亚、非欧等地区。**

The Company attaches great importance to brand awareness. While improving the ability and quality of management unceasingly, Jixiang people always uphold their creed of quality first, brand first. With unremitting efforts to occupy a leading position in China and even the world, the Company strives to build the reputation of Jixiang products both at home and abroad within two years.

**公司重视品牌意识，在不断提升公司管理水平、管理质量的基础上，坚持“质量第一，品牌至上”是集翔人永恒的信条，公司力争在二年内打响“集翔”牌产品在国内外外的知名度，为跻身全国乃至世界前列不懈努力。**

**福建集翔获得长汀政府大力扶持！**

# Fujian Jixiang Medical Technology Co., Ltd

福建集翔医疗科技有限公司



福建集翔获得长汀政府大力扶持！

# Fujian Jixiang Medical Technology Co., Ltd

福建集翔医疗科技有限公司



# Technical support

## 技术支持

Our company's equipment is configured by the domestic professional equipment research and development team, the daily capacity can reach 1.09 million, is the domestic production line with the highest daily capacity.

**我们公司设备由国内专业设备研发团队配置，日产能可达109万只，是国内目前日产能最高的生产线。**

The company's glove commissioning team is made up of engiAneer Zou, who has been engaged in glove commissioning for more than 20 years. Engineer Zou has been invited by dozens of glove factories at home and abroad to commissioning equipment

**公司手套调试团队由从业超20年的邹工程师为核心组成，邹工程师曾受到国内外数十家手套厂邀请到厂调试设备**



# 建线进度预估表 (战略合作伙伴合作新建线数量未列入)

福建集翔丁腈手套生产线生产计划表（三十九条生产线）						
设备编号	框架完成日期	设备安装完成	调试	生产	手套产量（每天万只）	手套产量（每月万只）
第一条线	9月20日	10月20日	10月30日	12月10日	109	3052
第二条线	9月20日	10月30日	11月10日	12月15日	218	6104
第三条线	9月30日	12月20日	12月30日	1月8日	327	9156
第四条线	9月30日	12月20日	12月30日	1月8日	436	12208
第五条线	10月10日	12月20日	12月30日	1月8日	545	15260
第六条线	10月10日	12月20日	12月30日	1月8日	654	18312
第七条线	10月10日	1月10日	1月20日	1月30日	763	21364
第八条线	10月10日	1月10日	1月20日	1月30日	872	24416
第九条线	11月20日	1月10日	1月20日	1月30日	981	27468
第十条线	11月20日	1月10日	1月20日	1月30日	1090	30520
第十一条线	11月20日	1月10日	1月20日	1月30日	1199	33572
第十二条线	11月20日	1月20日	1月30日	2月10日	1308	36624
第十三条线	11月20日	1月20日	1月30日	2月10日	1417	39676
第十四条线	11月20日	1月20日	1月30日	2月10日	1526	42728
第十五条线	11月20日	1月20日	1月30日	2月10日	1635	45780
第十六条线	11月20日	1月20日	1月30日	2月10日	1744	48832
第十七条线	12月30日	1月20日	1月30日	2月10日	1853	51884
第十八条线	12月30日	1月20日	1月30日	2月10日	1962	54936

# 建线进度预估表 (战略合作伙伴合作新建线数量未列入)

福建集翔丁腈手套生产线生产计划表（三十九条生产线）						
设备编号	框架完成日期	设备安装完成	调试	生产	手套产量（每天万只）	手套产量（每月万只）
第十九条线	12月30日	1月30日	2月20日	3月1日	2071	57988
第二十条线	12月30日	1月30日	2月20日	3月1日	2180	61040
第二十一条线	12月30日	1月30日	2月20日	3月1日	2289	64092
第二十二条线	12月30日	1月30日	2月20日	3月1日	2398	67144
第二十三条线	12月30日	1月30日	2月20日	3月1日	2507	70196
第二十四条线	12月30日	1月30日	2月20日	3月1日	2616	73248
第二十五条线	12月30日	1月30日	2月20日	3月1日	2725	76300
第二十六条线	12月30日	1月30日	2月20日	3月1日	2834	79352
第二十七条线	12月30日	1月30日	2月20日	3月1日	2943	82404
第二十八条线	12月30日	2月27日	3月25日	3月30日	3052	85456
第二十九条线	1月20日	2月27日	3月25日	3月30日	3161	88508
第三十条线	1月20日	2月27日	3月25日	3月30日	3270	91560
第三十一条线	1月20日	2月27日	3月25日	3月30日	3379	94612
第三十二条线	1月20日	2月27日	3月25日	3月30日	3488	97664
第三十三条线	1月20日	2月27日	3月25日	3月30日	3597	100716
第三十四条线	1月20日	2月27日	3月25日	3月30日	3706	103768
第三十五条线	1月20日	2月27日	3月25日	3月30日	3815	106820
第三十六条线	1月20日	2月27日	3月25日	3月30日	3924	109872
第三十七条线	1月20日	2月27日	3月25日	3月30日	4033	112924
第三十八条线	1月20日	2月27日	3月25日	3月30日	4142	115976
第三十九条线	1月20日	2月27日	3月25日	3月30日	4251	119028

# Certificate of approval of EIA

环评审评通过证明

编号:

**龙岩市生态环境局文件**

龙环审〔2020〕475号

**建设项目环境影响报告表**

项目名称: 年产70亿只医疗器械乳胶、丁腈手套生产线项目

建设单位: 福建集翔医疗科技有限公司 (盖章)

编制日期: 2020年10月

龙岩市生态环境局关于福建集翔医疗科技有限公司年产70亿只医疗器械乳胶、丁腈手套生产线项目环境影响报告表的批复

福建集翔医疗科技有限公司:

你单位《年产70亿只医疗器械乳胶、丁腈手套生产线项目环境影响报告表》(以下简称“报告表”)及申请审批的报告收悉。经研究,批复如下:

一、项目位于龙岩市长汀县工贸新城汀州大道南路29号,租用福建龙岩高新区长汀产业园区的长汀国有投资集团有限公司位于该处的现有厂房进行建设,总建筑面积78428m<sup>2</sup>,主要设置生产厂房,燃气锅炉和环保工程等。项目年产70亿只一类、二类医疗器械乳胶、丁腈手套,分两期建设,一期主要建设16

- 1 -

4

# WID ERANGE

OF PREMIUM QUILITY

优质产品 High quality products

# Examination gloves (Powder-free)

丁腈检查手套

无 粉



## QUALITY FEATURES

Biologically compatible  
Softness and shape provides excellent comfort  
Beaded cuff enhances donning and prevent rollback  
High tensile strength and elongation

## 质量特征：

生物相容性  
柔软性和准确形状提供了极好的舒适性  
珠状袖口增强了穿戴感，防止后卷高拉  
伸强度和伸长率

# Examination gloves (Powder-free)

丁腈检查手套 无粉



采用100%合成丁腈橡胶胶乳(丙烯腈丁二烯)制造与相同厚度的天然橡胶胶乳相比，可抵御多种化学品。

强度更高，抗穿刺性更好。

生物相容性。

柔软性和准确形状提供了极好的舒适性

珠状袖口增强了穿戴感，防止后卷

高拉伸强度和伸长率。

双手灵巧，手指可伸直。

无污垢痕迹、油污、嵌入的异物颗粒、凝结物等。完全不含滑石粉。

全纹理可抓握，在各种条件下都能获得出色的性能。

# Examination gloves (Powder-free)

丁腈检查手套 无粉



## PRODUCT SPECIFICATIONS

Material –Synthetic Nitrile Rubber Latex  
(Acrylonitrile Butadiene) Length –240 mm +/- 5mm  
Weight –4.0to 6.0g +/- 0.3g per piece Cuff Style  
–  
Beaded  
Surface –Palm Textured AQL –1- 2.5

## 产品规格

材料-合成丁腈橡胶胶乳(丙烯腈丁二烯)长度 –  
240mm+/-5mm  
重量-3.5至6.0克+/-0.3克/件袖口风格-珠状表面-棕  
榈纹理AQL-1-2.5



# Examination gloves polymer coated(Powder-free)

聚合物涂层乳胶检查手套

无 粉



## QUALITY FEATURES

Biologically compatible  
Softness and shape provides excellent comfort Beaded cuff enhances donning and prevent rollback High tensile strength and elongation  
Ambidextrous and straight fingers

## 质量特征：

生物相容性  
柔软性和准确形状提供了极好的舒适性  
珠状袖口增强了穿戴感，防止后卷高拉伸强度和伸长率

# Examination gloves polymer coated(Powder-free)

聚合物涂层乳胶检查手套

无 粉



## PRODUCT SPECIFICATIONS

Material – Synthetic Nitrile Rubber Latex (Acrylonitrile Butadiene) Length – 240 mm +/- 5mm  
Weight – 3.5 to 6.0g +/- 0.3g per piece Cuff Style – Beaded  
Surface – Palm Textured AQL – 1 - 2.5

## 产品规格

材料-合成丁腈橡胶胶乳(丙烯腈丁二烯)长度 – 240mm+/-5mm  
重量-3.5至6.0克+/-0.3克/件袖口风格-珠状表面-棕榈纹理AQL-1-2.5



# Examination gloves (Powder-free)

乳胶检查手套

无 粉



Manufactured with high quality and innovative manufacturing technology that makes the glove soft and provides excellent comfort. Beaded cuff enhances donning and prevents rollback.

High tensile strength and elongation. Ambidextrous and straight fingers. Free from dirt marks, oil stains, embedded foreign particles, coagulum etc. Extra washing and leaching reduces the possibility of allergic reactions to minimum level.

Polymer Coated for easy donning.

Fully textured grip for great performance in all conditions.

采用高质量创新制造技术制造，使手套柔软舒适。珠状袖口加强了穿戴和防止回卷。

高拉伸强度和伸长率。

双手灵巧，手指可伸直。

无污垢痕迹、油污、嵌入的异物颗粒、凝结物等。额外的洗涤和沥滤将过敏反应的可能性降至最低水平。聚合物涂层，便于穿戴。

全纹理可抓握，在各种条件下都能获得出色的性能。

# Examination gloves (Powder-free)

乳胶检查手套

无 粉



## QUALITY STANDARDS

Conforms to EN 455 standards.  
Manufactured under QRS (GMP) and ISO 9001:2015, ISO 14001:2015 Quality Management System CE certified and TUV Standard  
SGS Test report and FDA registration Hypoallergenic reducing potential allergic reactions. etc.

## 质量标准

符合EN455标准。  
根据QRS (GMP)和ISO 9001:2015、ISO 14001:2015质量管理体系CE认证和TUV标准制造根据QRS (GMP)和ISO 9001:2015、ISO 14001:SGS报告和美国食品和药物管理局注册。  
低过敏性，减少潜在的过敏反应。等等。

S, M, L, XL

# PVC gloves

聚氯乙烯手套



## 质量标准

- 符合EN455及EN374相关标准
- 符合ASTMD5250相关标准（美国市场产品）
- 符合EN455及EN374相关标准
- 符合ASTMD5250相关标准（美国市场产品）
- 拥有美国食药监局FDA510 (K)注册

## 产品特性

- 不含DOP及DEHP
- 无乳胶蛋白，相关过敏人士的良好选择
- 可用于食品处理，但油脂类食品除外
- 有粉及无粉产品可选
- 为多种物质提供良好防护
- 超软轻柔
- 具有良好手感，佩戴舒适
- AQL水平1.5 I 2.5 I 4,0可选

福建集翔聚氯乙烯手套有粉及无粉系列以100%全新改良配方的材料，带来比以往产品更为轻软的感受，同时确保对多种物质和微生物良好的防护。这是良好感受与物理特性的平衡，同时是对乳胶手套中蛋白过敏人士的良好替代方案。

# Synthetic gloves

混丁. 合成手套



## 质量标准

符合EN455及EN374相关标准  
符合ASTMD5250相关标准（美国市场产品）

## 产品特性


- 全新一代检查手套产品
- 不含DOP及DEHP
- 无天然乳胶蛋白，极低过敏风险
- 接近天然乳胶的舒适手感
- 全面的防护性与超强韧性
- 增强的敏锐手感
- 蓝色、白色、紫色粉色等多种流行色
- 最具性价比的手部保护方案

**革命性的第四代检查手套，全新的聚氯乙烯基复合材料确保了接近橡胶的弹性手感与强韧“骨感”。新材料带来全面的防护与敏锐手感，更是乳胶蛋白过敏人士的上佳之选。**

# OUR EXCLUSIVERANGE



# OUR EXCLUSIVERANGE



# OUR EXCLUSIVERANGE



# QUALITY ASSURANCE

## 质量保证

Product meets all started and implied needs. Manufactured under hygiene and environment control. Superior barrier protection. Scientifically tested gloves. Standardized and Customized Packing. Well equipped microbiology lab. State-of-the-art manufacturing inspection, and packing facilities.





产品满足所有已开始和隐含的需求。 在卫生和环境控制下制造。  
卓越的屏障保护。  
经过科学测试的手套。  
标准化和定制包装。  
设备完善的微生物实验室。  
最先进的制造检验和包装设施。




# Specifications

现有产品规格

产品品名	型号	净重 (KGS)	尺寸	单位	毛重 (千 克)	包装方 式	装箱尺寸	照片
丁腈手套	S	4.0	9英寸	只	5.3	100 只/ 盒  10 盒/ 箱	31.5x26.5x25.5cm	
	M	4.5	9英寸	只	5.85			
	L	5.0	9英寸	只	6.35		1300 箱/20 尺距	
	XL	5.5	9英寸	只	6.85		3300 箱/40 尺距	

产品品名	型号	净重 (KGS)	尺寸	单位	毛重 (千 克)	包装方 式	装箱尺寸	照片
PVC 手套	S	4.0	9英寸	只	4.85	100 只/ 盒  10 盒/ 箱	31.5x26.5x25.5cm	
	M	4.5	9英寸	只	5.35			
	L	5.0	9英寸	只	5.85		1300 箱/20 尺距	
	XL	5.5	9英寸	只	6.35		3300 箱/40 尺距	

产品品名	型号	净重 (KGS)	尺寸	只	毛重 (千 克)	包装方 式	装箱尺寸	照片
一次性混合 丁腈手 套	S	3.8	9英寸	只	4.65	100 只/ 盒  10 盒/ 箱	31.5x26.5x25.5cm	
	M	4.3	9英寸	只	5.15			
	L	4.8	9英寸	只	6.65		1300 箱/20 尺距	
	XL	5.3	9英寸	只	6.15		3300 箱/40 尺距	

产品品名	型号	净重 (KGS)	尺寸	只	毛重 (千 克)	包装方 式	装箱尺寸	照片
乳胶手套	S	5.0	9英寸	只	5.85	100 只/ 盒  10 盒/ 箱	31.5x26.5x25.5cm	
	M	5.5	9英寸	只	6.35			
	L	6.0	9英寸	只	6.85		1300 箱/20 尺距	
	XL	6.5	9英寸	只	7.35		3300 箱/40 尺距	

国标检测报告医用级丁腈 Nitrile GB 10213-2006

国标检测报告医用级丁腈 Nitrile GB 10213-2006

报告编号: WT202005076

第 1 页 共 1 页

说明

1. 报告不单独对委托送检检测项目进行说明。
2. 报告无指定人员签字无效。报告可修改、撤销无效。
3. 本检测标准依据书号编写, 不单独对复验本检测报告, 未检项目不作为报告广告使用。
4. 委托标准改造项目不在本检测单 CMA、CNAS 认可范围内, 本报告只对该检测样品检测结果负责。
5. 委托本单位对本检测标准有异议, 请在收到报告之日起按约定领取报告之日起, 7 个工作日内提出申诉, 逾期不予受理。
6. 如客户特别注明对样品保密要求, 所有超过标准规定时限的样品均不再保留样。

深圳市佳信检测服务有限公司

http://www.jiexin.com.cn

报告编号: JX\_202005076


检测地址: 深圳市宝安区西乡街道西乡社区西乡大道4号 检测: 0755-26666666 邮编: 518101

检测地址: 深圳市宝安区西乡街道西乡社区西乡大道4号 检测: 0755-26666666 邮编: 518101

总公司地址: 深圳市宝安区西乡街道 电话: 0755-26666666 邮编: 518101

# examining report butyronitrile

国标检测报告医用级丁腈 Nitrile GB 10213-2006



## 检验报告 TEST REPORT

报告编号: WT208020576

产品名称: 丁腈检查手套

委托单位: 福建集翔医疗科技有限公司

单位地址: 福建省龙岩市武平武平大道南路 29 号

生产单位: 福建集翔医疗科技有限公司

生产单位地址: 福建省龙岩市武平武平大道南路 29 号

检测日期: 2020 年 08 月 19 日-2020 年 09 月 15 日

检测类别: 委托检测

深圳市计量质量检测研究院 批准人: 邓海英

签发日期: 2020.09.16 签名: 邓海英

深圳市计量质量检测研究院  
地址: 深圳市福田区香蜜湖街道 4 号 总机: 0755-88291000 邮编: 518055  
业务受理部: 深圳市福田区香蜜湖街道 4 号 总机: 0755-88291000 邮编: 518055  
业务受理部: 深圳市福田区香蜜湖街道 4 号 总机: 0755-88291000 邮编: 518055

报告编号: WT208020576

第 1 页 共 5 页

## 检测报告

样品名称	丁腈检查手套	产品型号	2005
样品	1 号	规格	2020 年 08 月 19 日
检测项目	1. 外观检查 2. 尺寸检查 3. 物理性能 4. 化学性能 5. 卫生性能		
检测结果	符合 GB 10213-2006 要求		
检测日期	2020 年 08 月 19 日	检测地点	实验室
检测人员	邓海英	审核人员	邓海英

报告编号: WT208020576

第 2 页 共 5 页

## 检测报告

序号	检测项目	标准要求	检测结果	判定
1	物理性能	拉伸强度	12.32	合格
		断裂伸长率	11.28	合格
		撕裂强度	13.13	合格
		耐磨性能	13.13	合格
2	化学性能	耐油性	11.70	合格
		耐水性	11.70	合格
		耐汗液	11.70	合格
		耐酸碱	11.70	合格
3	卫生性能	细菌菌落总数	1.00	合格
		霉菌菌落总数	1.00	合格
		大肠杆菌	1.00	合格
		金黄色葡萄球菌	1.00	合格

产品照片



## 说明

1. 报告无防伪标志和防伪检测专用章无效。
2. 报告无签发人签名无效, 报告经修改、增删无效。
3. 本报告检测数据仅供参考, 不作为法律依据, 本报告不作为法律诉讼的依据。
4. 委托检测项目不在本实验室 CNAS、CMA 认可范围内, 本报告只对送检样品检测结果负责。
5. 委托本实验室检测, 请在收到报告之日起 10 个工作日内提出异议, 逾期不予受理。
6. 本报告不作为法律诉讼的依据, 本报告不作为法律诉讼的依据。

深圳市计量质量检测研究院  
地址: 深圳市福田区香蜜湖街道 4 号 总机: 0755-88291000 邮编: 518055  
业务受理部: 深圳市福田区香蜜湖街道 4 号 总机: 0755-88291000 邮编: 518055  
业务受理部: 深圳市福田区香蜜湖街道 4 号 总机: 0755-88291000 邮编: 518055

深圳市计量质量检测研究院  
地址: 深圳市福田区香蜜湖街道 4 号 总机: 0755-88291000 邮编: 518055  
业务受理部: 深圳市福田区香蜜湖街道 4 号 总机: 0755-88291000 邮编: 518055  
业务受理部: 深圳市福田区香蜜湖街道 4 号 总机: 0755-88291000 邮编: 518055

深圳市计量质量检测研究院  
地址: 深圳市福田区香蜜湖街道 4 号 总机: 0755-88291000 邮编: 518055  
业务受理部: 深圳市福田区香蜜湖街道 4 号 总机: 0755-88291000 邮编: 518055  
业务受理部: 深圳市福田区香蜜湖街道 4 号 总机: 0755-88291000 邮编: 518055

深圳市计量质量检测研究院  
地址: 深圳市福田区香蜜湖街道 4 号 总机: 0755-88291000 邮编: 518055  
业务受理部: 深圳市福田区香蜜湖街道 4 号 总机: 0755-88291000 邮编: 518055  
业务受理部: 深圳市福田区香蜜湖街道 4 号 总机: 0755-88291000 邮编: 518055

SGS检测报告医用级丁腈 EN 455

	
<b>Test Report</b>	No.: XMHL2010010152CW
	Date: Nov 02, 2020      Page 1 of 4
<b>FUJIAN JIXIANG MEDICAL TECHNOLOGY CO., LTD</b> 29 TINGZHOU AVENUE SOUTH ROAD, CEWU TOWN, CHANGTING COUNTY, FUJIAN PROVINCE	
Sample Description :	NITRILE GLOVES
As above test item and its relevant information regarding to the submission are provided and confirmed by the applicant. SGS is not liable to either the test item or its relevant information, in terms of the accuracy, suitability, reliability or/and integrity accordingly.	
SGS Ref No. :	QDHL2010010736MD
Sample Receiving Date :	Oct 10, 2020
Test Performing Date :	Oct 10, 2020 to Nov 02, 2020
Test Performed :	Selected test(s) as requested by applicant
Test Result(s) :	For further details, please refer to the following page(s)
Signed for and on behalf of SGS-CSTC Standards Technical Services Co., Ltd. Xiamen Branch	  
Beck Hong Authorized Signatory	
	
<p>I hereby acknowledge agreed in writing. This document is issued by the Company subject to its General Conditions of Service printed under separate cover or attached at the end of this report. Terms and conditions apply to all reports unless otherwise stated. The Company reserves the right to amend its services without notice. The Company's liability is limited to the amount of the fee paid for the service. The Company is not responsible for any loss or damage caused by the use of the service. The Company is not responsible for any loss or damage caused by the use of the service. The Company is not responsible for any loss or damage caused by the use of the service.</p> <p>SGS - Société Générale de Surveillance S.A. - 100 Rue du Rhône - 1201 Genève - Suisse          Tel: +41 (0)22 80 87 1000 Fax: +41 (0)22 80 87 1001 Email: info@sgs.com Web: www.sgs.com</p>	

[illegible]

# examining report PVC

SGS检测报告医用级PVC

GB 24786-2009

深圳市计量质量检测研究院  
Shenzhen Academy of Metrology & Quality Inspection

MA  
CNAS

### 检验报告

TEST REPORT

报告编号: WT208020575

产品名称: PVC 一次性手套

委托单位: 福建集翔医疗科技有限公司

单位地址: 福建省长汀县武江大道南路 29 号

生产单位: 福建集翔医疗科技有限公司

生产单位地址: 福建省长汀县武江大道南路 29 号

检测日期: 2020 年 08 月 19 日至 2020 年 09 月 15 日

检测类别: 委托检测

深圳市计量质量检测研究院 批准人: 邓海英  
签发日期: 2020.09.15 签名: 邓海英

深圳市计量质量检测研究院 <http://www.smg.org.cn> 电子邮箱: [sgs@smg.org.cn](mailto:sgs@smg.org.cn)  
客户服务热线: 0755-85000000 地址: 深圳市福田区香蜜湖 4 号 邮编: 518000  
总机: 0755-85000000 分机: 0755-85000000 传真: 0755-85000000  
客户服务热线: 0755-85000000 地址: 深圳市福田区香蜜湖 4 号 邮编: 518000

报告编号: WT208020575

### 检测报告

样品名称	规格	检测项目	检测结果
样品名称	规格	检测项目	检测结果
1. 外观	1. 外观	1. 外观	1. 外观
2. 物理性能	2. 物理性能	2. 物理性能	2. 物理性能
3. 化学性能	3. 化学性能	3. 化学性能	3. 化学性能
4. 力学性能	4. 力学性能	4. 力学性能	4. 力学性能
5. 卫生性能	5. 卫生性能	5. 卫生性能	5. 卫生性能
6. 其他	6. 其他	6. 其他	6. 其他

检测日期: 2020 年 08 月 19 日至 2020 年 09 月 15 日

检测类别: 委托检测

深圳市计量质量检测研究院 批准人: 邓海英  
签发日期: 2020.09.15 签名: 邓海英

深圳市计量质量检测研究院 <http://www.smg.org.cn> 电子邮箱: [sgs@smg.org.cn](mailto:sgs@smg.org.cn)  
客户服务热线: 0755-85000000 地址: 深圳市福田区香蜜湖 4 号 邮编: 518000  
总机: 0755-85000000 分机: 0755-85000000 传真: 0755-85000000  
客户服务热线: 0755-85000000 地址: 深圳市福田区香蜜湖 4 号 邮编: 518000

报告编号: WT208020575

### 检测报告

序号	检测项目	标准要求	检测结果	判定
1	外观	1. 外观	1. 外观	合格
2	物理性能	2. 物理性能	2. 物理性能	合格
3	化学性能	3. 化学性能	3. 化学性能	合格
4	力学性能	4. 力学性能	4. 力学性能	合格
5	卫生性能	5. 卫生性能	5. 卫生性能	合格
6	其他	6. 其他	6. 其他	合格

检测日期: 2020 年 08 月 19 日至 2020 年 09 月 15 日


检测类别: 委托检测

深圳市计量质量检测研究院 批准人: 邓海英  
签发日期: 2020.09.15 签名: 邓海英

深圳市计量质量检测研究院 <http://www.smg.org.cn> 电子邮箱: [sgs@smg.org.cn](mailto:sgs@smg.org.cn)  
客户服务热线: 0755-85000000 地址: 深圳市福田区香蜜湖 4 号 邮编: 518000  
总机: 0755-85000000 分机: 0755-85000000 传真: 0755-85000000  
客户服务热线: 0755-85000000 地址: 深圳市福田区香蜜湖 4 号 邮编: 518000

报告编号: WT208020575

### 产品照片



检测日期: 2020 年 08 月 19 日至 2020 年 09 月 15 日

检测类别: 委托检测

深圳市计量质量检测研究院 批准人: 邓海英  
签发日期: 2020.09.15 签名: 邓海英

深圳市计量质量检测研究院 <http://www.smg.org.cn> 电子邮箱: [sgs@smg.org.cn](mailto:sgs@smg.org.cn)  
客户服务热线: 0755-85000000 地址: 深圳市福田区香蜜湖 4 号 邮编: 518000  
总机: 0755-85000000 分机: 0755-85000000 传真: 0755-85000000  
客户服务热线: 0755-85000000 地址: 深圳市福田区香蜜湖 4 号 邮编: 518000

报告编号: WT208020575

### 说明

1. 报告无须经本检测和校准实验室盖章无效。

2. 报告无须经本检测和校准实验室盖章无效。

3. 本报告书仅对来样负责, 不对来样真实性负责, 本报告书不作为法律依据。

4. 本报告书仅对来样负责, 不对来样真实性负责, 本报告书不作为法律依据。

5. 本报告书仅对来样负责, 不对来样真实性负责, 本报告书不作为法律依据。

6. 本报告书仅对来样负责, 不对来样真实性负责, 本报告书不作为法律依据。

检测日期: 2020 年 08 月 19 日至 2020 年 09 月 15 日

检测类别: 委托检测

深圳市计量质量检测研究院 批准人: 邓海英  
签发日期: 2020.09.15 签名: 邓海英

深圳市计量质量检测研究院 <http://www.smg.org.cn> 电子邮箱: [sgs@smg.org.cn](mailto:sgs@smg.org.cn)  
客户服务热线: 0755-85000000 地址: 深圳市福田区香蜜湖 4 号 邮编: 518000  
总机: 0755-85000000 分机: 0755-85000000 传真: 0755-85000000  
客户服务热线: 0755-85000000 地址: 深圳市福田区香蜜湖 4 号 邮编: 518000

# examining report PVC

SGS检测报告医用级PVC EN 455

**SGS**

**Test Report** No.: XMHL2010010076CW Date: Oct 28, 2020 Page 1 of 4

FUJIAN JIXIANG MEDICAL TECHNOLOGY CO., LTD  
29 TINGZHOU AVENUE SOUTH ROAD, CEWU TOWN, CHANGTING COUNTY, FUJIAN PROVINCE

Sample Description : DISPOSABLE PVC GLOVES

As above test item and its relevant information regarding to the submission are provided and confirmed by the applicant. SGS is not liable to either the test item or its relevant information, in terms of the accuracy, suitability, reliability or/and integrity accordingly.

SGS Ref No.: QDHL2010010644MD

Sample Receiving Date : Oct 10, 2020

Test Performing Date : Oct 10, 2020 to Oct 28, 2020

Test Performed : Selected test(s) as requested by applicant

Test Result(s) : For further details, please refer to the following page(s)

Signed for and on behalf of  
SGS-CSTC Standards Technical Services Co., Ltd. Xiamen Branch

Beck Hong  
Authorized Signatory

Member of the SGS Group (SGS SA)

**SGS**

**Test Report** No.: XMHL2010010076CW Date: Oct 28, 2020 Page 2 of 4

Test Items	Water tightness test, Tensile strength (Force at break, Force at break after challenge testing), Removable surface powder, Proteins, leachable
Testing Accordance	EN 455-1:2020 Medical Gloves for Single Use – Part 1: Requirements and Testing for Freedom from Holes Clause 5.1 EN 455-2:2015 Medical Gloves for Single Use – Part 2: Requirements and Testing for Physical Properties Clause 5.2.5.3 EN 455-3:2015 Medical Gloves for Single Use – Part 3: Requirements and Testing for Biological Evaluation Clause 4.4.4.5

**Test Results**

Test Items	Unit	Test Method	Requirement	Test Result	Assessment
Water tightness test	l	EN455-1: 2020 Clause 5.1	Sample quantity: 200pcs AQL: 1.5 Ac: 7 Re: 8	Found: 0	Pass
Tensile strength	Force at break	N	EN 455-2: 2015 Clause 5.2	Median value: ≥3.6	Pass
	Force at break after challenge testing	N	EN 455-2: 2015 Clause 5.3	Median value: ≥3.6	Pass
Removable surface powder	mg	EN 455-3: 2015 Clause 4.4 EN ISO 21171:2006	≤2	0.10	Pass
Proteins, leachable	µg/g	EN 455-3: 2015 Clause 4.5	l	Not detected (Method Detection Limit: 10)	Pass

Member of the SGS Group (SGS SA)

**SGS**

**Test Report** No.: XMHL2010010076CW Date: Oct 28, 2020 Page 3 of 4



**Appendix 1: Tensile strength**

Size: M			
Force at break (N)			
Before aging		After aging	
No.	l	No.	l
1	4.3	1	3.6
2	3.6	2	3.8
3	4.0	3	3.7
4	4.0	4	4.1
5	4.0	5	3.5
6	4.2	6	3.8
7	4.8	7	4.3
8	4.4	8	3.7
9	4.0	9	4.5
10	3.9	10	3.6
11	4.0	11	3.6
12	4.1	12	4.2
13	4.3	13	4.0
Standard requirement	≥3.6	Standard requirement	≥3.6
Median value	4.0	Median value	3.8

Remark:  
1. The declaration of conformity is only based on the actual value of laboratory activity, measurement uncertainty of the results not take into account.  
2. This test was subcontracted to Qingdao Branch, SGS-CSTC Co., Ltd.

Member of the SGS Group (SGS SA)

# SGS检测报告医用级PVC EN 455



# examining report PVC

SGS检测报告医用级PVC EN 455

**SGS**

**Test Report** No.: XMHL20100101160T Date: Nov 02, 2020 Page 3 of 4

**Appendix 1: Dimensions**

Size	Length (mm)	Width (mm)
No.		
1	244	97
2	240	96
3	240	96
4	244	96
5	243	96
6	240	96
7	244	97
8	240	97
9	244	96
10	240	96
11	245	96
12	241	97
13	240	96
Standard requirement	≥240	≥94
Median value	241	96

**Appendix 2: Tensile strength**

Size: M			
Before aging		After aging	
No.	F	No.	F
1	5.5	1	5.6
2	5.2	2	5.2
3	5.4	3	5.2
4	5.6	4	5.3
5	5.5	5	5.1
6	5.0	6	5.2
7	5.0	7	5.1
8	5.6	8	5.4
9	5.4	9	5.0
10	5.2	10	5.0
11	5.0	11	5.1
12	5.2	12	5.2
13	5.5	13	5.2
Standard requirement	≥3.6	Standard requirement	≥3.6
Median value	5.4	Median value	5.2

**SGS**

SGS (Shanghai) Inspection & Certification Co., Ltd. (SGS SA) is a member of the SGS Group. The SGS Group is a leading provider of inspection, certification and testing services worldwide. The SGS Group is a member of the SGS Group (SGS SA).

**SGS**

**Test Report** No.: XMHL20100101160T Date: Nov 02, 2020 Page 4 of 4

**Remark:**

- The declaration of conformity is only based on the actual value of laboratory activity, measurement uncertainty of the results not take into account.
- This test was subcontracted to Qingdao Branch, SGS-CSTC Co., Ltd.

**Sample photo(s):**



The testing report/certificate only refers to the sample(s) tested.  
\*\*\*End of Report\*\*\*

**SGS**

SGS (Shanghai) Inspection & Certification Co., Ltd. (SGS SA) is a member of the SGS Group. The SGS Group is a leading provider of inspection, certification and testing services worldwide. The SGS Group is a member of the SGS Group (SGS SA).

# examining report PVC

国标检测报告食品级PVC

GB 4806.7-2016

深圳市计量质量检测研究院  
Shenzhen Institute of Metrology & Quality Inspection

**检验报告**  
TEST REPORT

报告编号: WT208020762

产品名称: PVC 一次性手套

委托单位: 福建集翔医疗科技有限公司

单位地址: 福建省龙岩市武平武平大道南路 29 号

生产单位: 福建集翔医疗科技有限公司

生产单位地址: 福建省龙岩市武平武平大道南路 29 号

检测日期: 2020 年 08 月 13 日-2020 年 09 月 15 日

检测类别: 委托检测

批准人: 邓海英

签发日期: 2020.09.15

签名: 邓海英

报告编号: WT208020762

**检测报告**

样品名称	样品规格	样品数量
样品名称	样品规格	样品数量

检测项目: 1. 感官指标; 2. 理化指标; 3. 卫生指标; 4. 安全指标; 5. 其他指标

检测方法: 1. 感官指标: 目视、鼻嗅、手触; 2. 理化指标: 气相色谱-质谱联用; 3. 卫生指标: 微生物检测; 4. 安全指标: 重金属检测; 5. 其他指标: 其他检测方法

检测结果: 1. 感官指标: 符合 GB 4806.7-2016 要求; 2. 理化指标: 符合 GB 4806.7-2016 要求; 3. 卫生指标: 符合 GB 4806.7-2016 要求; 4. 安全指标: 符合 GB 4806.7-2016 要求; 5. 其他指标: 符合 GB 4806.7-2016 要求

报告编号: WT208020762

**检测报告**

序号	检测项目	检测结果	判定
1	感官指标	符合 GB 4806.7-2016 要求	合格
2	理化指标	符合 GB 4806.7-2016 要求	合格
3	卫生指标	符合 GB 4806.7-2016 要求	合格
4	安全指标	符合 GB 4806.7-2016 要求	合格
5	其他指标	符合 GB 4806.7-2016 要求	合格

报告编号: WT208020762

**产品照片**



说明

1. 报告无防伪标识, 请客户自行鉴别真伪。  
2. 报告无防伪标识, 请客户自行鉴别真伪。  
3. 本报告仅对送检样品负责, 不作为其他用途。  
4. 本报告仅对送检样品负责, 不作为其他用途。  
5. 本报告仅对送检样品负责, 不作为其他用途。  
6. 本报告仅对送检样品负责, 不作为其他用途。

报告编号: WT208020762

**说明**

1. 报告无防伪标识, 请客户自行鉴别真伪。  
2. 报告无防伪标识, 请客户自行鉴别真伪。  
3. 本报告仅对送检样品负责, 不作为其他用途。  
4. 本报告仅对送检样品负责, 不作为其他用途。  
5. 本报告仅对送检样品负责, 不作为其他用途。  
6. 本报告仅对送检样品负责, 不作为其他用途。

## GB 10213-2006

报告编号: W288020642

第 2 页 共 2 页

## 说 明

- 1、符合无纸化录单和检验流程与绿通要求。
- 2、报告无签发人签名盖章，报告结论无效，请重新检测。
- 3、本送本检测机构书面问题，本海部分复印本检测报告，本检测报告不得作为商业广告使用。
- 4、带\*标准或项目不是本实验室 CMA、CNAS 认可范围内，本报告只对该检测样品检测负责。
- 5、委托本单位对本检测报告有异议，请在收到报告之日起或规定领取报告之日起，7个工作日内向本检测中心提出，逾期不予受理。
- 6、报告为特殊行业贸易文件材料品管使用，所有在检测标准测定时加测的样品均不另收费。

检测单位计量资质检测范围

http://www.cma.cn.cn

电子邮箱: [zj\\_cma@163.com](mailto:zj_cma@163.com)


水南无纸化系统: 检测项目地址: 检测地址: 4号 电话: 2755-8888888 邮编: 310003

北京无纸化系统: 检测项目地址: 检测地址: 4号 电话: 2755-8888888 邮编: 310003

金华无纸化系统: 检测项目地址: 检测地址: 4号 电话: 2755-8888888 邮编: 310003

# examining report Latex

SGS检测报告医用级乳胶 EN 455

**Test Report** No.: XMHL20100101160T Date: Nov 02, 2020 Page 1 of 4


FUJIAN JIXIANG MEDICAL TECHNOLOGY CO., LTD.  
29 TINGZHOU AVENUE SOUTH ROAD, CEWU TOWN, CHANGTING COUNTY, FUJIAN PROVINCE


Sample Description : LATEX GLOVES  
As above test item and its relevant information regarding to the submission are provided and confirmed by the applicant. SGS is not liable to either the test item or its relevant information, in terms of the accuracy, suitability, reliability or/and integrity accordingly.

SGS Ref No. : QDHL2010010737MD  
Sample Receiving Date : Oct 12, 2020  
Test Performing Date : Oct 12, 2020 to Nov 02, 2020  
Test Performed : Selected test(s) as requested by applicant  
Test Result(s) : For further details, please refer to the following page(s)

Signed for and on behalf of:  
SGS-CSTC Standards Technical Services Co., Ltd. Xiamen Branch


Beck Hong  
Authorized Signatory

  
Member of the SGS Group (SGS SA)

**Test Report** No.: XMHL20100101160T Date: Nov 02, 2020 Page 2 of 4

**Test Results**

Test Items	Unit	Test Method	Requirement	Test Result	Assessment
Water tightness test	/	EN455-1: 2020 Clause 5.1	Sample quantity: 200pcs AQL: 1.5 Ac: 7 Re: 8	Found: 5	Pass
Dimensions	Length	mm	EN 455-2: 2015 Clause 4.2	Median value: M: 2240	Pass
	Width	mm	EN 455-2: 2015 Clause 4.3	Median value: M: 95±10	Pass
Tensile strength	Force at break	N	EN 455-2: 2015 Clause 5.2	Median value: ≥3.6	Pass
	Force at break after challenge testing	N	EN 455-2: 2015 Clause 5.3	Median value: ≥3.6	Pass
Removable surface powder	mg	EN 455-3: 2015 Clause 4.4 EN ISO 21171:2006	≤2	0.32	Pass
Proteins, leachable	µg/g	EN 455-3: 2015 Clause 4.5	/	137 (Method Detection Limit: 10)	/

  
Member of the SGS Group (SGS SA)

SGS检测报告医用級乳胶 EN 455

No.	Size	M
	Length (mm)	Width (mm)
1	244	97
2	240	96
3	240	96
4	244	96
5	243	96
6	240	96
7	244	97
8	240	97
9	244	96
10	240	96
11	245	96
12	241	97
13	240	96
Standard requirement	≥240	95±10
Median value	241	96

**Appendix 2: Tensile strength**

Size: M			
Before aging		After aging	
No.	F	No.	F
1	5.5	1	5.6
2	5.2	2	5.2
3	5.4	3	5.2
4	5.6	4	5.3
5	5.5	5	5.1
6	5.0	6	5.2
7	5.0	7	5.1
8	5.6	8	5.4
9	5.4	9	5.0
10	5.2	10	5.0
11	5.0	11	5.1
12	5.2	12	5.2
13	5.5	13	5.2
Standard requirement	≥3.6	Standard requirement	≥3.6
Median value	5.4	Median value	5.2

**Appendix 3: Impact resistance**

Size: M			
Before aging		After aging	
No.	E	No.	E
1	1.0	1	1.0
2	1.0	2	1.0
3	1.0	3	1.0
4	1.0	4	1.0
5	1.0	5	1.0
6	1.0	6	1.0
7	1.0	7	1.0
8	1.0	8	1.0
9	1.0	9	1.0
10	1.0	10	1.0
11	1.0	11	1.0
12	1.0	12	1.0
13	1.0	13	1.0
Standard requirement	≤1.0	Standard requirement	≤1.0
Median value	1.0	Median value	1.0

**Appendix 4: Water absorption**

Size: M			
Before aging		After aging	
No.	G	No.	G
1	0.05	1	0.05
2	0.05	2	0.05
3	0.05	3	0.05
4	0.05	4	0.05
5	0.05	5	0.05
6	0.05	6	0.05
7	0.05	7	0.05
8	0.05	8	0.05
9	0.05	9	0.05
10	0.05	10	0.05
11	0.05	11	0.05
12	0.05	12	0.05
13	0.05	13	0.05
Standard requirement	≤0.05	Standard requirement	≤0.05
Median value	0.05	Median value	0.05

**Appendix 5: Thermal stability**

Size: M			
Before aging		After aging	
No.	T <sub>g</sub>	No.	T <sub>g</sub>
1	145	1	145
2	145	2	145
3	145	3	145
4	145	4	145
5	145	5	145
6	145	6	145
7	145	7	145
8	145	8	145
9	145	9	145
10	145	10	145
11	145	11	145
12	145	12	145
13	145	13	145
Standard requirement	140~150	Standard requirement	140~150
Median value	145	Median value	145

**Appendix 6: Dimensional change**

Size: M			
Before aging		After aging	
No.	L	No.	L
1	0.05	1	0.05
2	0.05	2	0.05
3	0.05	3	0.05
4	0.05	4	0.05
5	0.05	5	0.05
6	0.05	6	0.05
7	0.05	7	0.05
8	0.05	8	0.05
9	0.05	9	0.05
10	0.05	10	0.05
11	0.05	11	0.05
12	0.05	12	0.05
13	0.05	13	0.05
Standard requirement	≤0.05	Standard requirement	≤0.05
Median value	0.05	Median value	0.05

**Appendix 7: Surface roughness**

Size: M			
Before aging		After aging	
No.	Ra	No.	Ra
1	0.05	1	0.05
2	0.05	2	0.05
3	0.05	3	0.05
4	0.05	4	0.05
5	0.05	5	0.05
6	0.05	6	0.05
7	0.05	7	0.05
8			

## Test Report

No.: XMHL20100101160T

Date: Nov 02, 2020

Page 4 of 4

### Remark:

1. The declaration of conformity is only based on the actual value of laboratory activity, measurement uncertainty of the results not take into account.
2. This test was subcontracted to Qingsdao Branch, SGS-CSTC Co., Ltd.

### Sample photo(s):

The testing report/certificate only refers to the sample(s) tested.

\*\*\*End of Report\*\*\*

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed and available at [www.sgs.com](http://www.sgs.com) or [www.sgs.com.cn](http://www.sgs.com.cn). The Company's liability is limited to the amount of the fee paid for the service. The Company is not responsible for any damage or loss of any kind, including consequential loss, arising from the use of the results of the service. The Company is not responsible for any damage or loss of any kind, including consequential loss, arising from the use of the results of the service. The Company is not responsible for any damage or loss of any kind, including consequential loss, arising from the use of the results of the service.

SGS (Shanghai) Branch, 100000 Shanghai, China  
中国·上海·南京路100000号  
Tel: +86 21 5855 5788 / Fax: +86 21 5855 5789 / Email: [sgs@sgs.com.cn](mailto:sgs@sgs.com.cn)

Member of the SGS Group (SGS SA)

# examining report Synthetic

国标检测报告医用级混丁 Synthetic GB 10213-2006

深圳市计量质量检测研究院  
Shenzhen Academy of Metrology & Quality Inspection

IMA A

## 检验报告

TEST REPORT

报告编号: WT2091063726

产品名称: 一次性混合丁腈检查手套

委托单位: 福建集翔医疗科技有限公司

单位地址: 福建省省长汀县武镇汀州大道南路 29 号

生产单位: 福建集翔医疗科技有限公司

生产单位地址: 福建省省长汀县武镇汀州大道南路 29 号

检测日期: 2020 年 10 月 30 日-2020 年 11 月 06 日

检测类别: 委托检测

批准人: 邓海英

签发日期: 2020.09.15

签名: 邓海英

深圳市计量质量检测研究院 批准人: 邓海英

签发日期: 2020.09.15

签名: 邓海英

深圳市计量质量检测研究院 批准人: 邓海英

签发日期: 2020.09.15

签名: 邓海英

报告编号: WT2091063726 第 4 页, 共 4 页

## 检测报告

样品名称	一次混合丁腈检查手套	产品型号	2006
样品	10 瓶	检测日期	2020 年 10 月 30 日
检测标准	GB 10213-2006	检测项目	1. 外观检查 2. 尺寸检查 3. 厚度检查 4. 拉伸性能 5. 穿刺性能 6. 摩擦性能
检测结果	符合标准	检测日期	2020 年 10 月 30 日
检测单位	深圳市计量质量检测研究院	检测日期	2020 年 10 月 30 日

深圳市计量质量检测研究院 批准人: 邓海英

签发日期: 2020.09.15

签名: 邓海英

报告编号: WT2091063726 第 4 页, 共 4 页

## 检测报告

序号	检测项目	检测结果	判定
1	外观检查	符合标准	合格
2	尺寸检查	符合标准	合格
3	厚度检查	符合标准	合格
4	拉伸性能	符合标准	合格
5	穿刺性能	符合标准	合格
6	摩擦性能	符合标准	合格

深圳市计量质量检测研究院 批准人: 邓海英

签发日期: 2020.09.15

签名: 邓海英

报告编号: WT2091063726 第 4 页, 共 4 页

## 产品照片



说明

1. 报告无修改和检验检测专用章无效。

2. 报告无签发人签名无效, 报告校核无效, 报告无效。

3. 本报告仅供内部使用, 不得用于其他用途, 未经同意不得作为商业广告使用。

4. 本报告或报告内容不在本实验室 CMA、CNAS 认可范围内, 本报告对送检样品检测结果负责。

5. 委托本单位对本单位检测项目, 请在收到报告之日起或规定期限内, 7 个工作日内提出申诉, 逾期不予受理。

6. 对客户特别申明支付样品管理费, 所有超过标准规定时数的样品均不再收费。

深圳市计量质量检测研究院 批准人: 邓海英

签发日期: 2020.09.15

签名: 邓海英

报告编号: WT2091063726 第 4 页, 共 4 页

## 产品照片



说明

1. 报告无修改和检验检测专用章无效。

2. 报告无签发人签名无效, 报告校核无效, 报告无效。

3. 本报告仅供内部使用, 不得用于其他用途, 未经同意不得作为商业广告使用。

4. 本报告或报告内容不在本实验室 CMA、CNAS 认可范围内, 本报告对送检样品检测结果负责。

5. 委托本单位对本单位检测项目, 请在收到报告之日起或规定期限内, 7 个工作日内提出申诉, 逾期不予受理。

6. 对客户特别申明支付样品管理费, 所有超过标准规定时数的样品均不再收费。

深圳市计量质量检测研究院 批准人: 邓海英

签发日期: 2020.09.15

签名: 邓海英

# examining report Synthetic

国标检测报告医用级蓝色合成防护 Synthetic GB 10213-2006

深圳市计量质量检测研究院  
Shenzhen Academy of Metrology & Quality Inspection

检验报告  
TEST REPORT

报告编号: WT2091063354

产品名称: 蓝色合成防护手套

委托单位: 福建集翔医疗科技有限公司

单位地址: 福建省省长汀县武镇汀州大道南路 29 号

生产单位: 福建集翔医疗科技有限公司

生产单位地址: 福建省省长汀县武镇汀州大道南路 29 号

检测日期: 2020 年 10 月 29 日-2020 年 11 月 06 日

检测类别: 委托检测

批准人: 何月行

签发日期: 2020.11.06

签名: 何月行

深圳市计量质量检测研究院  
地址: 深圳市福田区福强路 1 号 总机: 0755-89329999 邮编: 518052  
龙岗实验基地: 深圳市龙岗区龙城大道 4 号 总机: 0755-89329999 邮编: 518052  
龙华实验基地: 深圳市龙华区龙观大道 4 号 总机: 0755-29411000 邮编: 518055  
光明实验基地: 深圳市光明区光明大道 114 号 总机: 0755-29411000 邮编: 518109

报告编号: WT2091063326 第 4 页, 共 4 页

检测报告

样品名称: 一次性使用丁腈手套 产品型号: J1805

样品 份数: 10 份 检测日期: 2020 年 10 月 30 日

检测依据: 判定依据: GB 10213-2006 一次性使用医用橡胶手套

检测项目: 1. 外观质量检查  
2. 尺寸和重量检查  
3. 老化后性能检查  
4. 老化后性能检查  
5. 尺寸和重量检查

样品描述: 样品来源: 送检样品

检测日期: 2020 年 10 月 30 日 检测地点: 本公司实验室

检测结论: 根据上述检测结果, 判定样品符合标准, 产品符合国家标准 GB 10213-2006 的要求。

检测日期: 2020 年 11 月 06 日

报告编号: WT2091063354 第 4 页, 共 5 页

检验报告

测试项目	测试方法	限量	测试结果	单项判定
颜色: 蓝色	GB 8466-1-2014	色牢度: 无异常	色牢度: 无异常	合格
气味: 无味	GB 18881-2016	无异味	无异味	合格
拉伸强度: 10N/100mm <sup>2</sup>	GB 10213-2006	≥10N/100mm <sup>2</sup>	12N/100mm <sup>2</sup>	合格
断裂伸长率: 200%	GB 10213-2006	≥200%	220%	合格
扯断伸长率: 200%	GB 10213-2006	≥200%	220%	合格
扯断伸长率: 200%	GB 10213-2006	≥200%	220%	合格

检测日期: 2020 年 10 月 30 日 检测地点: 本公司实验室

检测结论: 根据上述检测结果, 判定样品符合标准, 产品符合国家标准 GB 10213-2006 的要求。

检测日期: 2020 年 11 月 06 日

报告编号: WT2091063354 第 4 页, 共 5 页

产品照片



说明

1. 报告无异常和检验报告专用章无效。
2. 报告无签发人签名无效, 报告无效。
3. 未经本检测机构和报告, 不得部分复制本检测报告, 未经授权不得作为“鉴定”使用。
4. 本报告或项目与本公司实验室 CMA、CNAS 认可范围内, 本报告只对送检样品负责。
5. 委托单位对本检测报告有异议, 请在此检测报告之日起规定期限内提出, 7 个工作日内由本中心, 逾期不予受理。
6. 每客户特别申明支付材料管理费, 所有超过标准规定时限的样品均不退还材料。

深圳市计量质量检测研究院  
地址: 深圳市福田区福强路 1 号 总机: 0755-89329999 邮编: 518052  
龙岗实验基地: 深圳市龙岗区龙城大道 4 号 总机: 0755-89329999 邮编: 518052  
龙华实验基地: 深圳市龙华区龙观大道 4 号 总机: 0755-29411000 邮编: 518055  
光明实验基地: 深圳市光明区光明大道 114 号 总机: 0755-29411000 邮编: 518109

报告编号: WT2091063354 第 4 页, 共 5 页

说明

1. 报告无异常和检验报告专用章无效。
2. 报告无签发人签名无效, 报告无效。
3. 未经本检测机构和报告, 不得部分复制本检测报告, 未经授权不得作为“鉴定”使用。
4. 本报告或项目与本公司实验室 CMA、CNAS 认可范围内, 本报告只对送检样品负责。
5. 委托单位对本检测报告有异议, 请在此检测报告之日起规定期限内提出, 7 个工作日内由本中心, 逾期不予受理。
6. 每客户特别申明支付材料管理费, 所有超过标准规定时限的样品均不退还材料。

深圳市计量质量检测研究院  
地址: 深圳市福田区福强路 1 号 总机: 0755-89329999 邮编: 518052  
龙岗实验基地: 深圳市龙岗区龙城大道 4 号 总机: 0755-89329999 邮编: 518052  
龙华实验基地: 深圳市龙华区龙观大道 4 号 总机: 0755-29411000 邮编: 518055  
光明实验基地: 深圳市光明区光明大道 114 号 总机: 0755-29411000 邮编: 518109

# examining report Nitrile

## 美标检测报告丁腈手套 ASTM D6319

**SGS**  
Test Report No.: XMHL2011010760T Date: Dec 12, 2020 Page 1 of 4

FUJIAN JIXIANG MEDICAL TECHNOLOGY CO., LTD.  
29 TINGZHOU AVENUE SOUTH ROAD, CEVU TOWN, CHANGTING COUNTY, FUJIAN PROVINCE

Sample Description: NITRILE GLOVES  
As above test item and its relevant information regarding to the submission are provided and confirmed by the applicant. SGS is not liable to either the test item or its relevant information, in terms of the accuracy, suitability, reliability and integrity accordingly.

SGS Ref No.: QDHL2011012919AD  
Sample Receiving Date: Nov 23, 2020  
Test Performing Date: Nov 23, 2020 to Dec 12, 2020  
Test Performed: Selected test(s) as requested by applicant  
Test Result(s): For further details, please refer to the following page(s)

Test information	Test items
Testing Accordance	Freedom from holes, Physical dimensions (Length, Width, Thickness), Powder residue for powder free gloves ASTM D6319-19 Standard Specification for Nitrile Examination Gloves for Medical Application Clause 6.1.2, 6.1.3, 6.1.5
Test conclusion	This report only provides the test results and individual judgment, conclusion please see below pages.
Remark	Issue date: DEC.11,2020

Signed for and on behalf of  
SGS-CSTC Standards Technical Services Co., Ltd. Xiamen Branch

Back Hong  
Authorized Signatory

**SGS**  
Test Report No.: XMHL2011010760T Date: Dec 12, 2020 Page 2 of 4

Test Items	Unit	Test Method	Requirement	Test Result	Assessment
Performance Requirements					
Freedom from Holes	/	ASTM D6319-19 Clause 7.3	Sample quantity: 20/pair AQL: 2.5 Ac: 10 Re: 11	Found: 0	Pass
Physical dimensions	Length	mm	L: 2230		Pass
	Width	mm	L: 110±10		Pass
	Thickness-finger	mm	ASTM D6319-19 Clause 7.4	See appendix 1 for details	Pass
	Thickness-palm	mm	Median Value: <0.05		Pass
Powder Residue For Powder Free Gloves	mg	ASTM D6319-19 Clause 7.6	≤2.0	0.14	Pass

**SGS**  
Test Report No.: XMHL2011010760T Date: Dec 12, 2020 Page 3 of 4

Appendix 1: Physical dimensions

Sample No.	Length/mm	Width/mm	Size: L	
			Thickness-finger/mm	Thickness-palm/mm
1	235	113	0.088	0.071
2	234	112	0.093	0.071
3	236	113	0.091	0.093
4	234	112	0.093	0.064
5	236	112	0.081	0.065
6	233	112	0.074	0.091
7	237	113	0.099	0.099
8	237	112	0.065	0.068
9	234	113	0.063	0.093
10	235	113	0.064	0.064
11	235	112	0.068	0.059
12	237	112	0.079	0.092
13	239	113	0.092	0.090
Standard requirement	2230	110±10	≤0.05	≤0.05
Found	0	0	0	0



# examining report Latex

## 美标检测报告乳胶手套 ASTM D3578

**SGS**

**Test Report** No.: XMHL2011010730T Date: Dec 12, 2020 Page 1 of 4

FUJIAN JIXIANG MEDICAL TECHNOLOGY CO., LTD  
29 TINGZHOU AVENUE SOUTH ROAD, CEWU TOWN, CHANGHONG COUNTY, FUJIAN PROVINCE

Sample Description : LATEX GLOVES

As above test item and its relevant information regarding to the submission are provided and confirmed by the applicant, SGS is not liable to either the test item or its relevant information, in terms of the accuracy, suitability, reliability or/and integrity accordingly.

SGS Ref No : QDHL2011012816MD

Sample Receiving Date : Nov 23, 2020

Test Performing Date : Nov 23, 2020 to Dec 12, 2020

Test Performed : Selected test(s) as requested by applicant

Test Result(s) : For further details, please refer to the following page(s)

Test Information	Test Items	Freedom from holes, Physical dimensions (Length, Width, Thickness), Powder residue for powder free gloves
Testing Accordance	ASTM D 3578-19 Standard Specification for Rubber Examination Gloves Clause 7.1.2, 7.1.3, 7.1.5	
Test conclusion	This report only provides the test results and individual judgment, conclusion please see follow pages. Issue date: DEC 11, 2020	
Remark	/	

Signed for and on behalf of  
SGS-CSTC Standards Technical Services Co., Ltd. Xiamen Branch

  
Beck Hong  
Authorized Signatory



**SGS**

**Test Report** No.: XMHL2011010730T Date: Dec 12, 2020 Page 2 of 4

Test Items	Unit	Test Method	Requirement	Test Result	Assessment
Freedom from Holes	/	ASTM D 3578-19 Clause 8.3	Sample quantity: 200pcs AQL: 2.5 Ac: 10 Re: 11	Found: 1	Pass
Physical dimensions	Length	mm	ASTM D 3578-19 Clause 8.4	L: 230	Pass
	Width	mm	ASTM D 3578-19 Clause 8.4	L: 111±10	Pass
	Thickness-finger	mm	ASTM D 3578-19 Clause 8.4	Sample quantity: 10pcs AQL: 4.0 Ac: 1 Re: 2	Pass
	Thickness-palm	mm	ASTM D 3578-19 Clause 8.4	Median Value: 0.08 Median Value: 0.08	Pass
Powder Residue For Powder Free Gloves	mg	ASTM D 3578-19 Clause 8.6	≤2.0	0.84	Pass

Remark:  
1. The declaration of conformity is only based on the actual value of laboratory activity, measurement uncertainty of the results not take into account.  
2. This test was subcontracted to Qingdao Branch, SGS-CSTC Co., Ltd.



**SGS**

**Test Report** No.: XMHL2011010730T Date: Dec 12, 2020 Page 3 of 4

Appendix 1: Physical Dimensions

Sample No.	Length (mm)	Width (mm)	Size L	
			Thickness-finger	Thickness-palm
1	246	106	0.107	0.090
2	243	106	0.110	0.096
3	247	106	0.114	0.097
4	251	106	0.105	0.094
5	245	107	0.110	0.096
6	245	106	0.102	0.096
7	243	106	0.107	0.090
8	246	107	0.114	0.092
9	242	106	0.105	0.090
10	246	106	0.102	0.083
11	248	106	0.109	0.089
12	242	106	0.112	0.090
13	243	106	0.103	0.090
Standard requirement	≥230	111±10	≤0.08	≤0.08
Found	0	0	0	0

Remark:  
1. The declaration of conformity is only based on the actual value of laboratory activity, measurement uncertainty of the results not take into account.  
2. This test was subcontracted to Qingdao Branch, SGS-CSTC Co., Ltd.




**SGS**

**Test Report** No.: XMHL2011010730T Date: Dec 12, 2020 Page 4 of 4

Sample Photo



\*\*\*End of Report\*\*\*



# examining report PVC

## 美标检测报告PVC手套 ASTM D5250

**SGS**  
Test Report No.: XMHL20110110720T Date: Dec 12, 2020 Page 1 of 4

FUJIAN JIXIANG MEDICAL TECHNOLOGY CO., LTD.  
29 TINGZHOU AVENUE SOUTH ROAD, CEWU TOWN, CHANGTING COUNTY, FUJIAN PROVINCE

Sample Description : DISPOSABLE PVC GLOVES  
As above test item and its relevant information regarding to the submission are provided and confirmed by the applicant. SGS is not liable to either the test item or its relevant information, in terms of the accuracy, suitability, reliability or/and integrity accordingly.

SGS Ref No. : QDHL2011012815MD

Sample Receiving Date : Nov 23, 2020

Test Performing Date : Nov 23, 2020 to Dec 12, 2020

Test Performed : Selected test(s) as requested by applicant

Test Result(s) : For further details, please refer to the following page(s)

Test Information	Test Items	Physical property characteristics (Tensile strength & Ultimate elongation before aging, Tensile strength & Ultimate elongation after aging)
Testing Accordance	ASTM D 5250-19 Standard Specification for Poly (Vinyl Chloride) Gloves for Medical Application Clause 6.1.4	
Test conclusion	This report only provides the test results and individual judgment, conclusion please see follow pages.	
Remark	Issue date: DEC 11, 2020	

Signed for and on behalf of  
SGS-CSTC Standards Technical Services Co., Ltd. Xiamen Branch

Beck Hong  
Authorized Signatory



**SGS**  
Test Report No.: XMHL20110110720T Date: Dec 12, 2020 Page 2 of 4

Test Items	Unit	Test Method	Requirement	Test Result	Assessment
Performance Requirements					
Physical property characteristics	Before Aging	Tensile strength	MPa	≥11	Pass
		Ultimate Elongation	%	ASTM D 5250-19 Clause 7.5	
	After Aging	Tensile strength	MPa	≥11	
		Ultimate Elongation	%	≥300	

**SGS**  
Test Report No.: XMHL20110110720T Date: Dec 12, 2020 Page 3 of 4

### Appendix 1: Physical property characteristics

Before Aging				After Aging			
Sample No.	Tensile strength (MPa)	Ultimate Elongation (%)	Sample No.	Tensile strength (MPa)	Ultimate Elongation (%)	Sample No.	Tensile strength (MPa)
1	20.7	356	1	16.4	318	1	16.4
2	19.3	359	2	16.4	316	2	16.4
3	16.4	306	3	15.5	312	3	15.5
4	20.6	375	4	16.1	333	4	16.1
5	17.6	327	5	17.0	329	5	17.0
6	17.5	326	6	16.2	316	6	16.2
7	18.5	351	7	16.7	300	7	16.7
8	18.0	331	8	17.3	330	8	17.3
9	21.3	363	9	19.7	363	9	19.7
10	18.4	334	10	16.3	317	10	16.3
11	18.0	377	11	20.0	360	11	20.0
12	16.5	241	12	16.4	305	12	16.4
13	19.7	366	13	15.1	349	13	15.1
Standard requirement	≥11	≥300	Standard requirement	≥11	≥300	Standard requirement	≥11
Found	0	0	Found	0	0	Found	0

Remark:  
1. The declaration of conformity is only based on the actual value of laboratory activity, measurement uncertainty of the results not take into account.  
2. This test was subcontracted to Qinglian Branch, SGS-CSTC Co., Ltd.

**SGS**  
Test Report No.: XMHL20110110720T Date: Dec 12, 2020 Page 4 of 4



\*\*\*End of Report\*\*\*



福建集翔

Fujian Jixiang

# examining report Nitrile

EN ISO 21420 ISO 3742 丁腈



SATRA Technology Services (Singapore) Ltd  
Unit 110, Jinchuan Garden, Xiang  
Newtown South, Singapore City  
Singapore 119971  
Tel: 485 25 780 2280/2281  
Email: info@satraty.com

Customer details: Fujian Jixiang Medical Technology Co., Ltd.  
No. 29 Tingzhou Avenue,  
Cebu Town,  
Chengdu County,  
Fujian Province,  
China  
Your reference: MDJ01  
Date of report: 23 December 2020  
Samples received: 27 November 2020  
Date(s) work carried out: 1-18 December 2020

## TECHNICAL REPORT

Subject: EN ISO 21420: 2020 size & dexterity & innocuousness test, EN ISO 374-2: 2019 air leak and water leak test on Disposable Powder free Nitrile gloves referenced as MDJ01, colour: Blue, Purple, size: 6, 7, 8

### Conditions of issue:

This report may be forwarded to other parties provided that it is not changed in any way. It must not be published, for example by including it in advertisements, without the prior, written permission of SATRA.

Results given in this report refer only to the samples submitted for analysis and tested by SATRA. Comments are for guidance only.

A satisfactory test report in no way implies that the product tested is approved by SATRA and no warranty is given as to the performance of the product tested. SATRA shall not be liable for any subsequent loss or damage incurred by the client as a result of information supplied in the report.

The uncertainty of the results (data) in this report is based on a standard uncertainty multiplied by a coverage factor k=2, which provides a coverage probability of approximately 95%.

Report signed by: Gladys He  
Position: Technologist  
Department: China Testing

(Page 1 of 5)

Gladys He



## TECHNICAL REPORT

### WORK REQUESTED

Samples described as Disposable Powder free Nitrile gloves referenced as MDJ01, colour: Blue, Purple, size: 6, 7, 8 were received by SATRA on 27 November 2020 for testing in accordance with EN ISO 21420: 2020 and EN ISO 374-2: 2019.

### SAMPLE SUBMITTED



Samples described as Disposable Powder free Nitrile gloves referenced as MDJ01, colour: Blue, Purple, size: 6, 7, 8

### TESTING REQUESTED

EN ISO 21420: 2020 Clause 5.1 - Sizing and measurement of gloves  
EN ISO 21420: 2020 Clause 5.2 - Dexterity  
EN ISO 374-2: 2019 Clause 7.2 - Air leak  
EN ISO 374-2: 2019 Clause 7.3 - Water leak  
EN ISO 21420: 2020 Clause 4.2 - Innocuousness of protective gloves

### CONCLUSION

The samples described as Disposable Powder free Nitrile gloves referenced as MDJ01, colour: Blue, Purple, size: 6, 7, 8 were found to achieve the following results:

EN ISO 21420: 2020 Clause 5.1 - See below table  
EN ISO 21420: 2020 Clause 5.2 - Level 5  
EN ISO 374-2: 2019 Clause 7.2 - Pass  
EN ISO 374-2: 2019 Clause 7.3 - Pass  
EN ISO 21420: 2020 Clause 4.2 - Pass PAHs, DMF and pH value

Detailed results are included on the following page(s)

Fujian Jixiang Medical Technology Co., Ltd.  
SATRA Reference: CHT030551/0248  
Date: 23 December 2020

Signature: Gladys He  
China Testing



## TECHNICAL REPORT

### Testing

Testing was carried out in accordance with EN ISO 21420: 2020 and EN ISO 374-2: 2019

Samples for testing were conditioned for at least 24 hours in a conditioned environment maintained at (23±2) °C and (50±5) % relative humidity.

### Requirements

Table 1 - Requirements for EN ISO 21420: 2020 Clause 5.2 Dexterity

Performance level	1	2	3	4	5
Diameter of dexterity pin mm	11.0	9.5	8.0	6.5	5.0

Table 2 - Requirements for EN ISO 374-2: 2019

Clause 7.2 Air leak	No leak to be detected
Clause 7.3 Water leak	No leak to be detected

### Test Results

Table 3 - EN ISO 21420: 2020 Test Results

Clause / Test	Requirement	Test Results	Unit (See note 4)	Result
5.1 Glove length, comfort and fit	N/A	Size	1 2 3	x 1.10 mm
		6	241 240 240	
		7	238 238 242	
		8	241 238 238	
5.2 Dexterity	See table 1	Minimum pin diameter / mm		N/A
		6	5.0	
		7	5.0	
		8	5.0	

Fujian Jixiang Medical Technology Co., Ltd.  
SATRA Reference: CHT030551/0248  
Date: 23 December 2020

(Page 3 of 5)

Signature: Gladys He  
China Testing



## TECHNICAL REPORT

Table 4 - EN ISO 374-2: 2019 Test Results

Clause / Test	Test Results	Unit (See note 4)	Result
7.2 Air leak test	Total air pressure used	2.9 kPa	N/A
	Sample size		
	6	No leaks detected	
	7	No leaks detected	
7.3 Water leak test	Sample size		N/A
	6	No leaks detected	
	7	No leaks detected	
	8	No leaks detected	

### Additional Information / Notes

Note 4 - Estimated uncertainty of measurement applied at point of test (e.g. to applied force or to tolerance limits) to ensure product meets requirements of the standard

Fujian Jixiang Medical Technology Co., Ltd.  
SATRA Reference: CHT030551/0248  
Date: 23 December 2020

(Page 4 of 5)

Signature: Gladys He  
China Testing



## TECHNICAL REPORT

### Innocuousness Test Results

Testing was conducted at a third-party laboratory and reported under their reference A201130003001. The laboratory is CNAS accredited to ISO 17025: 2017.

Sample Item	Sample Description	Location	Style
001	Disposable Powder free Nitrile gloves referenced as MDJ01, colour: Blue	Gloves	1
002	Disposable Powder free Nitrile gloves referenced as MDJ01, colour: Purple	Gloves	2

### pH Value - EN ISO 21420: 2020

Test Method 1: With reference to EN ISO 4045: 2018, analysed by pH meter.

Test Method 6: With reference to ISO 3071: 2005, analysed by pH meter.

Requirement:	1.5-8.5
--------------	---------

Test Item(s)	Unit	Result
Test Method	6	6
Parameter		
pH value of Extracting Solution	6.43	6.45
Temp. of Aqueous Extract	25.2	25.2
pH value of Aqueous Extract	6.1	6.2
Conclusion	PASS	PASS

Note / Key: deg. C = degree Celsius (°C) Temp. = Temperature

Remarks: Results (see notes) reported the average value from two trials.

Tested part(s) was/were specified by client.

Fujian Jixiang Medical Technology Co., Ltd.  
SATRA Reference: CHT030551/0248  
Date: 23 December 2020

(Page 5 of 5)

Signature: Gladys He  
China Testing

## EN ISO21420 ISO374 丁腈

EN ISO21420 ISO374 丁腈



# TECHNICAL REPORT

## Polycyclic Aromatic Hydrocarbons (PAHs) Content - EN ISO 21420:2013

Test Method: With reference to test method PD CEN ISO/TS 16190:2013

Maximum Allowable Limit	Each of all listed PAHs: 1.0 mg/kg
-------------------------	------------------------------------

Tested Item(s)	Result		Conclusion
	Detected Analyte(s)	Conc. Unit	
E001 + E002	ND	ND mg/kg	PASS

Note 1/Key ND = Not detected/(Detection Limit); Detection Limit (mg/kg): Each: 0.2; mg/kg = milligram per kilogram = ppm = part per million

Remark: The list of polycyclic aromatic hydrocarbons is summarized in table of Appendix. Tested particle size were specified by client. Composite testings) were performed by client.

### APPENDIX


List of Polycyclic Aromatic Hydrocarbons				
No.	Name of Analyte	CAS-No.	CAS-No.	
1	Chrysene	218-01-9	5 Dibenz (a, j) anthracene	55-76-3
2	Benzo (a) pyrene	56-54-4	6 Benzo (b) fluoranthene	253-26-3
3	Benzo (k) pyrene	192-67-2	7 Benzo (g) fluoranthene	255-43-3
4	Benzo (a) anthracene	56-55-3	8 Benzo (a) fluoranthene	257-48-9

Fujian Jiaxing Medical Technology Co., Ltd.  
SATRA Reference: CHT2005561 2068  
Date: 23 December 2020

(Page 6 of 8)

Signed

*Chen Ting*  
Chen Ting



# TECHNICAL REPORT

**Densify/Remanade/DMP/Al Content - EN ISO 21420:2002**

**Test Method:** With reference to EN 16778:2016, and then analyzed by Gas Chromatograph Mass Spectrometer

Analyte	Unit	Result	Client's Requirement
		Test Item(s)	
Densify/Remanade(DMP/Al)	mg/kg	001+ 002	1000
Conclusion	-	PASS	✓

**Note / Key:** ND = Not detected / Detection Limit      Detection Limit (mg/kg) : 5

mg/kg = milligram per kilogram = ppm = part per million

Tested article was/were specified by client.

Complete test(s) was/were specified by client.

\*\*\* End of Report \*\*\*

Fujian Jiang Medical Technology Co., Ltd.  
SATRA Reference: CHT0305581 2048  
Date: 23 December 2020

(Page 7 of 8)

Signed

*Calvin He*  
Quality Manager

[illegible][illegible]

# examining report Nitrile

EN ISO21420 ISO374 -3 丁腈

**SATRA**  
TECHNOLOGY

SATRA Technology Centre Ltd  
Worthington Way, Tisbury, Wiltshire,  
Northamptonshire, NN16 8SD, United Kingdom  
Tel: +44(0) 1208 410000  
Fax: +44(0) 1208 410000  
Email: info@satra.com  
www.satra.com

**Customer details:** SATRA Technology Services (Dongguan) Ltd SATRA reference: CHM0309602049LHA  
Unit 110, Xinhongmiao Garden  
Hongmiao Road  
Xiping, Hanchuan District  
DONGGUAN CITY  
Guangdong Province  
China  
523079

**Your reference:** CHT0000551  
**Date of report:** 13<sup>th</sup> January 2021  
**Samples received:** 30<sup>th</sup> November 2020  
**Date(s) work carried out:** 16<sup>th</sup> December 2020 to 11<sup>th</sup> January 2021

**TECHNICAL REPORT**

**SATRA Technology Services (Dongguan) Ltd:**  
Customer: Fujian Jixiang Medical Technology Co., Ltd.  
No. 28 Tianhu Avenue,  
Caoxi Town,  
Changling County,  
Fujian Province,  
China

**Subject:** EN 16523-1:2015+A1:2018 resistance to permeation by chemicals on gloves described as Disposable Powder free Nitrile gloves referenced as MDJ01, color: Blue, Purple

**Conditions of issue:**  
This report may be forwarded to other parties provided that it is not changed in any way. It must not be published, for example by including it in advertisements, without the prior written permission of SATRA.  
Results given in this report refer only to the samples submitted for analysis and tested by SATRA. Comments are for guidance only.  
Tests marked \* fall outside the UKAS Accreditation Schedule for SATRA. All interpretations of results of such tests and the comments based upon them are outside the scope of UKAS accreditation and are based on current SATRA knowledge.  
A satisfactory test report in no way implies that the product tested is approved by SATRA and no warranty is given as to the performance of the product tested. SATRA shall not be liable for any subsequent loss or damage incurred by the client as a result of information supplied in the report.  
The uncertainty of the results (k=2) in this report is based on a standard uncertainty multiplied by a coverage factor k=2, which provides a coverage probability of approximately 95%.

Report signed by: Lorna Harding  
Position: Technical  
Chemical & Analytical Technology  
(Page 1 of 8)

SATRA Technology Centre Ltd is a subsidiary of SATRA. Registered in England No. 388038 at the record address.

**SATRA**  
TECHNOLOGY

**TECHNICAL REPORT**

**WORK REQUESTED:**  
Samples of gloves described as Disposable Powder free Nitrile gloves referenced as MDJ01, color: Blue, Purple were received on the 30<sup>th</sup> November 2020 for testing in accordance with EN 16523-1:2015+A1:2018 and assessment in accordance with the requirements of EN ISO 374-1:2016+A1:2018.

**SAMPLES SUBMITTED:**

 Samples described as Disposable Powder free Nitrile gloves referenced as MDJ01, color: Blue

 Samples described as Disposable Powder free Nitrile gloves referenced as MDJ01, color: Purple

**CONCLUSION:**  
When assessed in accordance with the requirements of EN ISO 374-1:2016+A1:2018 the samples of gloves described as Disposable Powder free Nitrile gloves referenced as MDJ01, color: Blue, Purple achieved the following performance levels:

Chemical	Performance level
40% Sodium hydroxide (CAS: 1310-73-2)	6
30% Hydrogen peroxide (CAS: 7722-84-1)	1
37% Formaldehyde (CAS: 50-00-0)	5

Full results are reported in the following tables.

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0309602049LHA  
Date: 13<sup>th</sup> January 2021 (Page 2 of 8)

Signed: 

**SATRA**  
TECHNOLOGY

**TECHNICAL REPORT**

**TESTING REQUIRED:**  
• EN 16523-1:2015+A1:2018 - Determination of material resistance to permeation by chemicals - Part 1: Permeation by liquid chemical under conditions of continuous contact

**RESULTS AND REQUIREMENTS:**  
EN ISO 374-1:2016+A1:2018 - Protective gloves against dangerous chemicals and micro-organisms - Part 1: Terminology and performance requirements for chemical risks. Table 1: Permeation performance levels.

Permeation performance level	Measured breakthrough time (minutes)
1	>10
2	>30
3	>60
4	>120
5	>240
6	>480

Performance levels are based on the lowest individual result achieved per chemical.

EN 16523-1:2015+A1:2018 - Determination of material resistance to permeation by chemicals - Part 1: Permeation by liquid chemical under conditions of continuous contact

EN ISO 374-1:2016+A1:2018 - Protective gloves against dangerous chemicals and micro-organisms - Part 1: Terminology and performance requirements for chemical risks. Table 1: Permeation performance levels.

Using PTFE permeation cells with standardised dimensions

Specimen	Thickness (mm)	Breakthrough time (mins)
1 (Blue)	0.08	>480
2 (Purple)	0.06	>480
3 (Blue)	0.08	>480

Test result: **>480**  
Unit: **<t**

Visual appearance of specimens after testing: Swollen and discoloured

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0309602049LHA  
Date: 13<sup>th</sup> January 2021 (Page 3 of 8)

Signed: 

**SATRA**  
TECHNOLOGY

**TECHNICAL REPORT**

**Test/Property** Sample reference: Disposable Powder free Nitrile gloves referenced as MDJ01, color: Blue, Purple  
Chemical: 40% Sodium hydroxide  
Normalised permeation rate (NPR): 1 g/cm<sup>2</sup>/min  
Detection technique: Conductivity  
Collection medium: Deionised water (closed loop)  
Collection medium stirring rate: 45 - 65 min/min  
Test temperature: (23 ± 1) °C

**Performance** Level 6

EN 16523-1:2015+A1:2018 in accordance with SATRA SOP CAT-009


Using PTFE permeation cells with standardised dimensions

Specimen	Thickness (mm)	Breakthrough time (mins)
1 (Blue)	0.08	>480
2 (Purple)	0.06	>480
3 (Blue)	0.08	>480


Test result: **>480**  
Unit: **<t**

Visual appearance of specimens after testing: Swollen and discoloured

**PERMEATION RATE GRAPHS**  
Challenge chemical: 40% Sodium hydroxide



SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0309602049LHA  
Date: 13<sup>th</sup> January 2021 (Page 4 of 8)

Signed: 

**SATRA**  
TECHNOLOGY

**TECHNICAL REPORT**

**Test/Property** Sample reference: Disposable Powder free Nitrile gloves referenced as MDJ01, color: Blue, Purple  
Chemical: 30% Hydrogen peroxide  
Normalised permeation rate (NPR): 1 g/cm<sup>2</sup>/min  
Detection technique: Electrochemical detector  
Collection medium: Deionised water (closed loop)  
Collection medium stirring rate: 45 - 65 min/min  
Test temperature: (23 ± 1) °C

**Performance** Level 1

EN 16523-1:2015+A1:2018 in accordance with SATRA SOP CAT-025

Using PTFE permeation cells with standardised dimensions


Specimen	Thickness (mm)	Breakthrough time (mins)
1 (Purple)	0.07	Between 21 to 30
2 (Blue)	0.09	Between 46 to 60
3 (Purple)	0.07	Between 21 to 30

Test result: **Between 21 to 30**  
Unit: **See below**

Visual appearance of specimens after testing: Swollen and discoloured


For SOP CAT-025, where both the P1 and P2 are observed in the same sampling range, uncertainty is expressed as the time difference between the mid-point of the range and the previous sampling time. This uncertainty is included in the reported result.

**PERMEATION RATE GRAPHS**  
Challenge chemical: 30% Hydrogen Peroxide




Hydrogen peroxide is determined by discrete sampling, therefore the permeation rate graph is not a smooth curve.

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0309602049LHA  
Date: 13<sup>th</sup> January 2021 (Page 5 of 8)

Signed: 

# EN ISO21420 ISO374-3 丁腈



## EN ISO21420 ISO374 -4 丁腈



**SATRA Technology Services (Dongguan) Ltd:**  
Customer Fujian Jiaxing Medical Technology Co., Ltd  
No. 29 Tingzhou Avenue,  
Cesuo Town,  
Changting County,  
Fujian Province,  
China

**Subject:** EN ISO 374-4:2019 determination of resistance to degradation by dangerous chemicals on gloves described as Disposable Powder free Nitrile gloves referenced as MD301, color: Blue, Purple.

This report may be forwarded to other parties provided that it is not changed in any way. It must not be published, for example by including it in advertisements, without the prior, written permission of SATRA.

Results given in this report refer only to the samples submitted for analysis and tested by SATRA. Comments are for guidance only.

Tests marked # fall outside the UKAS Accreditation Schedule for SATRA. All interpretations of results of such tests are the comments based upon them are outside the scope of UKAS accreditation and are based on current SATRA knowledge.

A satisfactory test report in no way implies that the product tested is approved by SATRA and no warranty is given as to the performance of the product tested. SATRA shall not be liable for any subsequent loss or damage incurred by the client as a result of information supplied in the report.

The uncertainty of the results (UoM) in this report is based on a standard uncertainty multiplied by a coverage factor  $k$  which provides a coverage probability of approximately 95%.

Report signed by: Lorna Harding  
Position: Technologist  
Department: Chemical & Analytical Technology

(Page 1 of 5)

SA7004, Supplemental Document 1, is an addition to SA7004.1. Supplemental Document 1, SA7004.004, is the above, with



Samples described as Disposable Powder free Nitrile gloves reserved as MD101, color: Blue

Samples described as Disposable Powder free Nitrile gloves referenced as MD 101, color: Blue

When assessed in accordance with EN ISO 374-4:2019 the samples of gloves described as Disposable Powder free Nitrile gloves referenced as MDJ01, color: Blue, Purple achieved the following degradation results:


Chemical	Mean degradation / %
40% Sodium hydroxide (CAS: 1310-73-2)	-62.9
37% Formaldehyde (CAS: 50-00-0)	-43.9
30% Methylene chloride (CAS: 7243-84-4)	-3.3

- EN ISO 374-4:2019. Protective gloves against dangerous chemicals and micro-organisms. Part 4: Determination of resistance to degradation by chemicals

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM005690/2049LH  
Date: 13<sup>th</sup> January 2021

Signatures

(PAGE 2 OF 5)



Sample description:	Disposable Powder free Nitrile gloves referenced as MDJ01, color: Blue, Purple		
Challenge chemical:	40% Sodium hydroxide (CAS: 1310-73-2)		
Test temperature / °C:	(23 ± 1)		
Degradation / %:	Glove 1	Glove 2	Glove 3
Mean degradation (DR) / %:	-57.3	-63.4	-48.0
Standard deviation (σ <sub>DR</sub> ) / %:		62.9	
UoM / ± %:		18.3	
Appearance of samples after testing:		10.0	
		Slightly swollen	

<b>Sample description:</b>	Disposable Powder free Nitrile gloves referenced as MDJ01, color: Blue- Purple		
<b>Challenge chemical:</b>	37% Formaldehyde (CAS: 50-00-0)		
<b>Test temperature / °C:</b>	(23 ± 1)		
<b>Degradation / %:</b>	<b>Glove 1</b>	<b>Glove 2</b>	<b>Glove 3</b>
<b>Mean degradation (DR) / %:</b>	-68.4	-78.2	-5.9
<b>Standard deviation (σ<sub>95</sub>) / %:</b>		-43.9	
<b>UoM / ± %:</b>		36.8	
<b>Appearance of samples after testing:</b>		10.2	
		Slightly swollen	

<b>Sample description:</b>	Disposable Powder free Nitrile gloves referenced as MDJ01, color: Blue, Purple		
<b>Challenge chemical:</b>	30% Hydrogen peroxide (CAS: 7722-84-1)		
<b>Test temperature / °C:</b>	(23 ± 1)		
<b>Degradation %:</b>	<b>Glove 1</b>	<b>Glove 2</b>	<b>Glove 3</b>
	9.4	-15.5	0.2
<b>Mean degradation (DR) %:</b>		-2.3	
<b>Standard deviation (n%)</b>		13.1	
<b>UoM ± s %:</b>		11.1	
<b>Appearance of samples after testing:</b>	Stitching weakens and discoloured		

**NOTE:** Where the test specimens gave an increased puncture force after chemical exposure, the result is reported as a negative degradation.

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM03050002049/LH/B  
Date: 13<sup>th</sup> January 2021 (Page 3 of 5)

---

198



SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0305690/2049/LHB  
Date: 13<sup>th</sup> January 2021 (Page 4 of 6)

the quality of

---



SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0305690/2049/LH/B  
Date: 13th January 2024 (Page 5 of 5)

# examining report PVC

EN ISO21420 ISO374 PVC



SATRA Technology Services (Company) Ltd  
Unit 111, Ansheng Garden, Xiang  
Nanhai District, Dapeng City  
Guangdong Province, China  
Tel: +86 (0) 755 22660077  
Email: info@satra.com.cn

Customer details: Fujian Jixiang Medical Technology Co., Ltd.  
No. 25 Tongshu Avenue,  
Cenxi Town,  
Changting County,  
Fujian Province,  
China  
SATRA reference: CHT0305550 0048  
Your reference: MPVC01  
Date of report: 23 December 2020  
Samples received: 27 November 2020  
Date(s) work carried out: 01-18 December 2020

## TECHNICAL REPORT

Subject: EN ISO 21420: 2020 size & dexterity & innocuousness test, EN ISO 374-2: 2019 air leak and water leak test on Disposable Powder free PVC gloves referenced as MPVC01, colour: White, size: 6, 7, 8

### Conditions of Issue

This report may be forwarded to other parties provided that it is not changed in any way. It must not be published, for example by including it in advertisements, without the prior, written permission of SATRA.

Results given in this report refer only to the samples submitted for analysis and tested by SATRA. Comments are for guidance only.

A satisfactory test report in no way implies that the product tested is approved by SATRA and no warranty is given as to the performance of the product tested. SATRA shall not be liable for any subsequent loss or damage incurred by the client as a result of information supplied in this report.

The uncertainty of the results (UAM) in this report is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , which provides a coverage probability of approximately 95%.

Report signed by: Gladys He  
Position: Technologist  
Department: China Testing

(Page 1 of 9)

Gladys He



## TECHNICAL REPORT

### WORK REQUESTED

Samples described as Disposable Powder free PVC gloves referenced as MPVC01, colour: White, size: 6, 7, 8 were received by SATRA on 27 November 2020 for testing in accordance with EN ISO 21420: 2020 and EN ISO 374-2: 2019.

### SAMPLE SUBMITTED



Samples described as Disposable Powder free PVC gloves referenced as MPVC01, colour: White, size: 6, 7, 8

### TESTING REQUESTED

EN ISO 21420: 2020 Clause 5.1 – Sizing and measurement of gloves  
EN ISO 21420: 2020 Clause 5.2 – Dexterity  
EN ISO 374-2: 2019 Clause 7.2 – Air leak  
EN ISO 374-2: 2019 Clause 7.3 – Water leak  
EN ISO 21420: 2020 Clause 4.2 – Innocuousness of protective gloves

### CONCLUSION

The samples described as Disposable Powder free PVC gloves referenced as MPVC01, colour: White, size: 6, 7, 8 were found to achieve the following results:

EN ISO 21420: 2020 Clause 5.1 – See below table  
EN ISO 21420: 2020 Clause 5.2 – Level 5  
EN ISO 374-2: 2019 Clause 7.2 – Pass  
EN ISO 374-2: 2019 Clause 7.3 – Pass  
EN ISO 21420: 2020 Clause 4.2 – Pass PAHs, pH value, Phthalates and DMF

Detailed results are included on the following page(s)

Fujian Jixiang Medical Technology Co., Ltd.  
SATRA Reference: CHT0305550 0048  
Date: 23 December 2020

(Page 2 of 9)

Signature: Gladys He  
China Testing



## TECHNICAL REPORT

### Testing

Testing was carried out in accordance with EN ISO 21420: 2020 and EN ISO 374-2: 2019

Samples for testing were conditioned for at least 24 hours in a conditioned environment maintained at (23±2) °C and (50±5) % relative humidity.

### Requirements

Table 1 – Requirements for EN ISO 21420: 2020 Clause 5.2 Dexterity

Performance level	1	2	3	4	5
Diameter of dexterity pin item	11.0	9.5	8.0	6.5	5.0

Table 2 – Requirements for EN ISO 374-2: 2019

Clause 7.2 Air leak	No leak to be detected
Clause 7.3 Water leak	No leak to be detected

### Test Results

Table 3 – EN ISO 21420: 2020 Test Results

Clause / Test	Requirement	Test Results	UAM (See note 4)	Result
5.1 Glove length, comfort and fit	N/A	Size	Length: mm	x 1.10 mm
		1	2	
		6	238	
		237	239	
5.2 Dexterity	See table 1	Comfortable on fit	9	N/A
		239	238	
		235	236	
		234	234	
5.2 Dexterity	See table 1	Minimum pin diameter / mm	Size	Level 5
		6	5.0	
		7	5.0	
		8	5.0	

Fujian Jixiang Medical Technology Co., Ltd.  
SATRA Reference: CHT0305550 0048  
Date: 23 December 2020

(Page 3 of 9)

Signature: Gladys He  
China Testing



## TECHNICAL REPORT

Table 4 – EN ISO 374-2: 2019 Test Results

Clause / Test	Test Results	UAM (See note 4)	Result
7.2 Air leak test	Total air pressure used	2.9 kPa	N/A
	Sample size	Leak	
	6	No leaks detected	
	7	No leaks detected	
7.3 Water leak test	Sample size	Leak	N/A
	6	No leaks detected	
	7	No leaks detected	
	8	No leaks detected	

### Additional Information / Notes

Note 4 – Estimated uncertainty of measurement applied at point of test (e.g. to applied force or to tolerance limits) to ensure product meets requirements of the standard

Fujian Jixiang Medical Technology Co., Ltd.  
SATRA Reference: CHT0305550 0048  
Date: 23 December 2020

(Page 4 of 9)

Signature: Gladys He  
China Testing



## TECHNICAL REPORT

### Innocuousness Test Results

Testing was conducted at a third-party laboratory and reported under their reference A201130062001. The laboratory is CNAS accredited to ISO 17025: 2017

Sample Item	Sample Description	Location	Style
801	Disposable Powder free PVC gloves referenced as MPVC01, colour: White	Gloves	-

at Value: EN ISO 21420: 2020

Test Method 1 With reference to EN ISO 4045:2016, analysed by pH meter.  
Test Method 2 With reference to ISO 3071:2005, analysed by pH meter.

Requirement:	5.5.5	
Test Item	Unit	Result
Test Method	-	801
Parameter	-	8
pH Value of Extracting Solution	-	5.45
Temp. of Exposure Extract	deg. C	25.2
pH Value of Exposure Extract	-	7.8
Difference Figure	-	-
Conclusion	-	PASS

Note / Key: deg. C = degree Celsius (°C) Temp. = Temperature  
Remark: Result(s) was (were) reported the average value from two trials.  
Tested parties were/were notified by client

Fujian Jixiang Medical Technology Co., Ltd.  
SATRA Reference: CHT0305550 0048  
Date: 23 December 2020

(Page 5 of 9)

Signature: Gladys He  
China Testing








# examining report PVC




EN ISO21420 ISO374 -3 PVC

<b>SATRA</b> TECHNOLOGY	SATRA Technology Centre Ltd Wyndham Way, Telford Wre, Shropshire Telford Shropshire, SAT8 8DS, United Kingdom Tel +44 (0) 1828 610000 Fax +44 (0) 1828 610001 email: info@satra.com www.satra.com	
Customer details: SATRA Technology Services (Dongguan) Ltd Unit 115, Xunsheng Garden Hongwei Road Xiang, Nanhai District DONGGUAN CITY Guangdong Province China 523078	SATRA reference: CHM030689/0494/LC Your reference: JH Date of report: 19 <sup>th</sup> January 2021 Samples received: 30 <sup>th</sup> November 2020 Date(s) work carried out: 16 <sup>th</sup> December 2020 to 19 <sup>th</sup> January 2021	
<b>TECHNICAL REPORT</b>		
SATRA Technology Services (Dongguan) Ltd: Customer: Fujian Jixiang Medical Technology Co., Ltd. No. 28 Tingting Avenue, Caozuo Town, Changping County, Fujian Province, China		
Subject: EN 16523-1:2015+A1:2018 resistance to permeation by chemicals on gloves described as Disposable Powder free PVC gloves referenced as MPVC01		
<b>Conditions of Issue:</b> This report may be forwarded to other parties provided that it is not changed in any way. It must not be published, for example by including it in advertisements, without the prior written permission of SATRA. Results given in this report refer only to the samples submitted for analysis and tested by SATRA. Comments are for guidance only. Tests marked as full outside the UKAS Accreditation Schemes for SATRA. All interpretations of results of such tests and the comments based upon them are outside the scope of UKAS accreditation and are based on current SATRA knowledge. A satisfactory test report in no way implies that the product tested is approved by SATRA and no warranty is given as to the performance of the product tested. SATRA shall not be liable for any subsequent loss or damage incurred by the client as a result of information reported in this report. The uncertainty of the results (U <sub>95</sub> ) in this report is based on a standard uncertainty multiplied by a coverage factor k=2, which provides a coverage probability of approximately 95%.		
Report signed by: Position: Department:	Lilly Cove Technologist Chemical & Analytical Technology (Page 1 of 9)	
SATRA Technology Centre Ltd is a subsidiary of SATRA. Registration No. 399208 at the above address.		


<b>SATRA</b> TECHNOLOGY	<b>TECHNICAL REPORT</b>									
<b>WORK REQUESTED:</b> Samples of gloves described as Disposable Powder free PVC gloves referenced as MPVC01 were received on the 30 <sup>th</sup> November 2020 for testing in accordance with EN 16523-1:2015+A1:2018 and assessment in accordance with the requirements of EN ISO 374-1:2016+A1:2018.										
<b>SAMPLES SUBMITTED:</b> 										
Samples described as Disposable Powder free PVC gloves referenced as MPVC01										
<b>CONCLUSION:</b> When assessed in accordance with the requirements of EN ISO 374-1:2016+A1:2018 the samples of gloves described as Disposable Powder free PVC gloves referenced as MPVC01 achieved the following performance levels:										
<table><thead><tr><th>Chemical</th><th>Performance level</th></tr></thead><tbody><tr><td>40% Sodium hydroxide (CAS: 1310-73-2)</td><td>6</td></tr><tr><td>30% Hydrogen peroxide (CAS: 7722-84-1)</td><td>5</td></tr><tr><td>37% Formaldehyde (CAS: 50-09-5)</td><td>5</td></tr></tbody></table>	Chemical	Performance level	40% Sodium hydroxide (CAS: 1310-73-2)	6	30% Hydrogen peroxide (CAS: 7722-84-1)	5	37% Formaldehyde (CAS: 50-09-5)	5		
Chemical	Performance level									
40% Sodium hydroxide (CAS: 1310-73-2)	6									
30% Hydrogen peroxide (CAS: 7722-84-1)	5									
37% Formaldehyde (CAS: 50-09-5)	5									
Full results are reported in the following tables.										
SATRA Technology Services (Dongguan) Ltd SATRA Reference: CHM030689/0494/LC/A Date: 19 <sup>th</sup> January 2021 (Page 2 of 9)										
Signed: 										

<b>SATRA</b> TECHNOLOGY	<b>TECHNICAL REPORT</b>															
<b>TESTING REQUEST:</b> EN 16523-1:2015+A1:2018 - Determination of material resistance to permeation by chemicals - Part 1: Permeation by liquid chemical under conditions of continuous contact																
<b>RESULTS AND REQUIREMENTS:</b> EN ISO 374-1:2016+A1:2018 - Protective gloves against dangerous chemicals and micro-organisms - Part 1: Terminology and performance requirements for chemical risks. Table 1: Permeation performance levels.																
<table><thead><tr><th>Permeation performance level</th><th>Measured breakthrough time (minutes)</th></tr></thead><tbody><tr><td>1</td><td>&gt;10</td></tr><tr><td>2</td><td>&gt;30</td></tr><tr><td>3</td><td>&gt;60</td></tr><tr><td>4</td><td>&gt;120</td></tr><tr><td>5</td><td>&gt;240</td></tr><tr><td>6</td><td>&gt;480</td></tr></tbody></table>	Permeation performance level	Measured breakthrough time (minutes)	1	>10	2	>30	3	>60	4	>120	5	>240	6	>480		
Permeation performance level	Measured breakthrough time (minutes)															
1	>10															
2	>30															
3	>60															
4	>120															
5	>240															
6	>480															
Performance levels are based on the lowest individual result achieved per chemical.																
SATRA Technology Services (Dongguan) Ltd SATRA Reference: CHM030689/0494/LC/A Date: 19 <sup>th</sup> January 2021 (Page 3 of 9)																
Signed: 																

<b>SATRA</b> TECHNOLOGY	<b>TECHNICAL REPORT</b>													
<b>Test/Property</b> EN 16523-1:2015+A1:2018 in accordance with SATRA SCOP CAT-009														
<b>Sample reference:</b> Disposable Powder free PVC gloves referenced as MPVC01														
<b>Test information:</b> Chemical: 40% Sodium hydroxide Normalised permeation rate (NPR): 1 µg/cm <sup>2</sup> /min Detection technique: Conductivity (continuous measurement) Collection medium: Deionised water (closed loop) Collection medium stirring rate: 45 – 65 min/min Leak test protocol: within 4.02s Test temperature: (23 ± 1) °C														
<b>Performance:</b> Level 6														
<table><thead><tr><th>Specimen</th><th>Thickness (mm)</th><th>Breakthrough time (min)</th></tr></thead><tbody><tr><td>1</td><td>0.07</td><td>&gt;480</td></tr><tr><td>2</td><td>0.07</td><td>&gt;480</td></tr><tr><td>3</td><td>0.07</td><td>&gt;480</td></tr></tbody></table>	Specimen	Thickness (mm)	Breakthrough time (min)	1	0.07	>480	2	0.07	>480	3	0.07	>480		
Specimen	Thickness (mm)	Breakthrough time (min)												
1	0.07	>480												
2	0.07	>480												
3	0.07	>480												
<b>Test result:</b> U <sub>95</sub> >480														
Visual appearance of specimens after testing: Discoloured														
<b>PERMEATION RATE GRAPHS</b> Challenge chemical: 40% Sodium Hydroxide 														
SATRA Technology Services (Dongguan) Ltd SATRA Reference: CHM030689/0494/LC/A Date: 19 <sup>th</sup> January 2021 (Page 4 of 9)														
Signed: 														

<b>SATRA</b> TECHNOLOGY	<b>TECHNICAL REPORT</b>													
<b>Test/Property</b> EN 16523-1:2015+A1:2018 in accordance with SATRA SCOP CAT-009														
<b>Sample reference:</b> Disposable Powder free PVC gloves referenced as MPVC01														
<b>Test information:</b> Chemical: 30% Hydrogen peroxide Normalised permeation rate (NPR): 1 µg/cm <sup>2</sup> /min Detection technique: Electrochemical detector (periodic measurement) Collection medium: Deionised water (closed loop) Collection medium stirring rate: 45 – 65 min/min Leak test protocol: within 4.02s Test temperature: (23 ± 1) °C														
<b>Performance:</b> Level 5														
<table><thead><tr><th>Specimen</th><th>Thickness (mm)</th><th>Breakthrough time (min)</th></tr></thead><tbody><tr><td>1</td><td>0.08</td><td>&gt;480</td></tr><tr><td>2</td><td>0.07</td><td>&gt;480</td></tr><tr><td>3</td><td>0.07</td><td>&gt;480</td></tr></tbody></table>	Specimen	Thickness (mm)	Breakthrough time (min)	1	0.08	>480	2	0.07	>480	3	0.07	>480		
Specimen	Thickness (mm)	Breakthrough time (min)												
1	0.08	>480												
2	0.07	>480												
3	0.07	>480												
<b>Test result:</b> U <sub>95</sub> Between 361 to 480														
Visual appearance of specimens after testing: Swollen and discoloured														
<b>PERMEATION RATE GRAPHS</b> Challenge chemical: 30% Hydrogen Peroxide 														
For SCOP CAT-025, where both the P <sub>i</sub> and P <sub>u</sub> are observed in the same sampling range, uncertainty is expressed as the time difference between the mid-point of the range and the previous sampling time. This uncertainty is included in the reported result.														
Hydrogen peroxide is determined by discrete sampling; therefore the permeation rate graph is not a smooth curve.														
SATRA Technology Services (Dongguan) Ltd SATRA Reference: CHM030689/0494/LC/A Date: 19 <sup>th</sup> January 2021 (Page 5 of 9)														
Signed: 														

# EN ISO21420 ISO374 -3 PVC




**PERMEATION RATE GRAPHS**  
Challenge chemical: 37% Formaldehyde

Y-axis: Average Permeation rate (gpm/m²)  
X-axis: Time / min

Legend:  
 - Specimen 1 (Blue line)  
 - Specimen 2 (Red line)  
 - Specimen 3 (Green line)  
 - 37% Formaldehyde (Dashed blue line)

Time (min)	Specimen 1 (gpm/m²)	Specimen 2 (gpm/m²)	Specimen 3 (gpm/m²)
0	0.0	0.0	0.0
30	0.1	0.1	0.1
60	0.2	0.2	0.2
90	0.3	0.3	0.3
120	0.4	0.4	0.4
150	0.5	0.5	0.5
180	0.6	0.6	0.6
210	0.7	0.7	0.7
240	0.8	0.8	0.8
270	0.9	0.9	0.9
300	1.0	1.0	1.0
330	1.1	1.1	1.1
360	1.2	1.2	1.2
390	1.3	1.3	1.3
420	1.3	1.3	1.3
450	1.3	1.3	1.3
480	1.3	1.3	1.3

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0305689/20491/C/A  
Date: 15<sup>th</sup> January 2021 (Page 6 of 9)



SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0305669/2049/LC/A  
Date: 15<sup>th</sup> January 2021 (Page 7 of 9)



SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0305689/2048LC/A  
Order: 1670, Amount: 30000 (RMB 0 of 0)



SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0305669/2049/LC/A  
Date: 10th January 2021 (Page 9 of 9)

Signed: *[Signature]*

# EN ISO21420 ISO374-4 PVC

	SATRA Technology Centre Ltd Wytham Way, Telford, Shropshire, North Worcestershire, WV10 6JG, United Kingdom Tel: +44 (0) 1824 40000 Fax: +44 (0) 1824 40001 email: info@satra.com www.satra.com		  CECE
	SATRA Technology Services (Dongguan) Ltd Unit 115, Zhongxing Garden Hongxing Road Xiang, Nanchang District DONGGUAN CITY Guangdong Province China 523079		
Customer details:	Your reference: Date of report: Samples received: Date(s) work carried out:	SATRA reference: CHM305068920483.C CHT0305050 10 <sup>th</sup> January 2021 30 <sup>th</sup> November 2020 8 <sup>th</sup> to 13 <sup>th</sup> January 2021	



# TECHNICAL REPORT



---

**WORK REQUESTED:**

Samples of gloves described as Disposable Powder free PVC gloves referenced as MPVC01 were received on the 30<sup>th</sup> November 2020 for testing in accordance with EN ISO 374-4:2019.

**SAMPLE SUBMITTED:**

**TESTED:**

**TESTED BY:**

**TESTED ON:**

**SATRA** TECHNOLOGY

**TECHNICAL REPORT**

**RESULTS:**

Sample description:	Dispositive Powder free PVC gloves referenced at MP-021		
Challenge chemical:	40% Sodium hydroxide (CAS: 1310-73-2)		
Test temperature °C:	(23 ± 1)		
Degradation / %:	Glove 1	Glove 2	Glove 3
	5.9	-2.9	0.0

[illegible][illegible]

# examining report Latex

EN ISO21420 ISO374 乳胶



SATRA Technology Services (Singapore) Ltd  
Unit 111, Ardchangan Garden, Xiang  
Nanyang Road, Singapore City  
Singapore 630001, China  
Tel: +65 6740 2200  
Email: info@satra.sg

Customer details: Fujian Jixiang Medical Technology Co., Ltd.  
No. 29 Tangzhou Avenue,  
Cebu Town,  
Changting County,  
Fujian Province,  
China

SATRA reference: CHT0305549 0048

Your reference: MRJ01

Date of report: 23 December 2020

Samples received: 27 November 2020

Date(s) work carried out: 01-18 December 2020

## TECHNICAL REPORT

Subject: EN ISO 21420: 2020 size & dexterity & innocuousness test, EN ISO 374-2: 2019 air leak and water leak test on Disposable Powdered Latex Gloves referenced as MRJ01, Colour: White, size: 6, 7

### Conditions of Issue:

This report may be forwarded to other parties provided that it is not changed in any way. It must not be published, for example by including it in advertisements, without the prior written permission of SATRA.

Results given in this report refer only to the samples submitted for analysis and tested by SATRA. Comments are for guidance only.

A satisfactory test report in no way implies that the product tested is approved by SATRA and no warranty is given as to the performance of the product tested. SATRA shall not be liable for any subsequent loss or damage incurred by the client as a result of information supplied in the report.

The uncertainty of the results (U<sub>95</sub>) in this report is based on a standard uncertainty multiplied by a coverage factor k=2, which provides a coverage probability of approximately 95%.

Report signed by: Gladys He  
Position: Technologist  
Department: China Testing

(Page 1 of 5)



## TECHNICAL REPORT

### WORK REQUESTED

Samples described as Disposable Powdered Latex Gloves referenced as MRJ01, Colour: White, size: 6, 7 were received by SATRA on 27 November 2020 for testing in accordance with EN ISO 21420: 2020 and EN ISO 374-2: 2019.

### SAMPLE SUBMITTED



Samples described as Disposable Powdered Latex Gloves referenced as MRJ01, Colour: White, size: 6, 7

### TESTING REQUESTED

EN ISO 21420: 2020 Clause 5.1 – Sizing and measurement of gloves

EN ISO 21420: 2020 Clause 5.2 – Dexterity

EN ISO 374-2: 2019 Clause 7.2 – Air leak

EN ISO 374-2: 2019 Clause 7.3 – Water leak

EN ISO 21420: 2020 Clause 4.2 – Innocuousness of protective gloves

### CONCLUSION

The samples described as Disposable Powdered Latex Gloves referenced as MRJ01, Colour: White, size: 6, 7 were found to achieve the following results:

EN ISO 21420: 2020 Clause 5.1 – See below table

EN ISO 21420: 2020 Clause 5.2 – Level 5

EN ISO 374-2: 2019 Clause 7.2 – Pass

EN ISO 374-2: 2019 Clause 7.3 – Pass

EN ISO 21420: 2020 Clause 4.2 – Pass PAVs and pH value

Detailed results are included on the following page(s)

Fujian Jixiang Medical Technology Co., Ltd.  
SATRA Reference: CHT0305549 0048  
Date: 23 December 2020

(Page 2 of 5)

Signature: Gladys He  
Position: Technologist  
Department: China Testing



## TECHNICAL REPORT

### Testing

Testing was carried out in accordance with EN ISO 21420: 2020 and EN ISO 374-2: 2019

Samples for testing were conditioned for at least 24 hours in a conditioned environment maintained at (23±2) °C and 55±5% relative humidity.

### Requirements

Table 1 – Requirements for EN ISO 21420: 2020 Clause 5.2 Dexterity

Performance level	1	2	3	4	5
Diameter of dexterity pin /mm	11.0	9.5	8.0	6.5	5.0

Table 2 – Requirements for EN ISO 374-2: 2019

Clause 7.2 Air leak	No leak to be detected
Clause 7.3 Water leak	No leak to be detected

### Test Results

Table 3 – EN ISO 21420: 2020 Test Results

Clause / Test	Requirement	Test Results	U <sub>95</sub> (See note 4)	Result
5.1 Glove length, comfort and fit	N/A	Size	1 2 3	N/A
		Length/mm	6 245 245 247	
5.2 Dexterity	See table 1	Comfortable on fit	7 245 245 245	Level 5
		Comfortable on fit	7 245 245 245	
5.2 Dexterity	See table 1	Size	6 5.0	N/A
		Minimum pin diameter / mm	6 5.0	
		7 5.0	7 5.0	
		7 5.0	7 5.0	

Fujian Jixiang Medical Technology Co., Ltd.  
SATRA Reference: CHT0305549 0048  
Date: 23 December 2020

(Page 3 of 5)

Signature: Gladys He  
Position: Technologist  
Department: China Testing



## TECHNICAL REPORT

Table 4 – EN ISO 374-2: 2019 Test Results

Clause / Test	Test Results	U <sub>95</sub> (See note 4)	Result
7.2 Air leak test	Total air pressure used	3.0 kPa	N/A
	Sample size	Leak	
	6	No leaks detected	
	6	No leaks detected	
7.3 Water leak test	Sample size	Leak	N/A
	6	No leaks detected	
	7	No leaks detected	
	7	No leaks detected	

### Additional Information / Notes

Note 4 – Estimated uncertainty of measurement applied at point of test (e.g. to applied force or to tolerance limit) to ensure product meets requirements of the standard



## TECHNICAL REPORT

### Innocuousness Test Results

Testing was conducted at a third-party laboratory and reported under their reference A201130061001. The laboratory is CNAS accredited to ISO 17025: 2017

Sample Item	Sample Description	Location	Style
001	Disposable Powdered Latex Gloves referenced as MRJ01, Colour: White	Gloves	-

### pH Value – EN ISO 21420: 2020

Test Method 1: With reference to EN ISO 4045:2016, analysed by pH meter

Test Method 2: With reference to ISO 3071:2020, analysed by pH meter

Requirement	3.5-5.5
Link	Result
Link	8
Test Method	8
Parameter	5.5
pH Value of Expiration Solution	5.5
Temp. of Exposure Extract	23.2
pH Value of Exposure Extract	5.3
Difference Factor	0.1
Conclusion	PASS

Note / Key: deg. C = degree Celsius (°C) Temp. = Temperature

Remark: Results were tested against the average value from two trials.

Tested part(s) was/were specified by client.

Fujian Jixiang Medical Technology Co., Ltd.  
SATRA Reference: CHT0305549 0048  
Date: 23 December 2020

(Page 5 of 5)

Signature: Gladys He  
Position: Technologist  
Department: China Testing



# examining report Latex

EN ISO21420 ISO374 -3 乳胶

**SATRA TECHNOLOGY**

SATRA Technology Centre Ltd  
Windsor Way, Telford Way, Redditch  
Worcestershire, B11 1SD, United Kingdom  
Tel: +44 (0) 1827 810000  
Fax: +44 (0) 1827 410008  
email: info@satra.com  
www.satra.com

**Customer details:** SATRA Technology Services (Dongguan) Ltd SATRA reference: CHM030568/2048/LC  
Unit 115, Xingcheng Garden  
Hongwei Road  
Xingcheng District  
DONGGUAN CITY  
Guangdong Province  
China  
523375

**Yours reference:** CHT000549  
**Date of report:** 12<sup>th</sup> January 2021  
**Samples received:** 30<sup>th</sup> November 2020  
**Date(s) work carried out:** 16<sup>th</sup> December 2020 to 12<sup>th</sup> January 2021

**TECHNICAL REPORT**

**SATRA Technology Services (Dongguan) Ltd:**  
Customer: Fujian Jixiang Medical Technology Co., Ltd.  
No. 28 Taigong Avenue,  
Caoxi Town,  
Changping County,  
Fujian Province,  
China

**Subject:** EN 15523-1:2015+A1:2018 resistance to permeation by chemicals on gloves described as Disposable Powdered Latex Gloves referenced as MRJ01

**Conditions of issue:**  
This report may be forwarded to other parties provided that it is not changed in any way, it must not be published, for example by including it in advertisements, without the prior, written permission of SATRA.  
Results given in this report refer only to the samples submitted for analysis and tested by SATRA. Comments are for guidance only.  
Tests marked # fall outside the UKAS Accreditation Schedule for SATRA. All interpretations of results of such tests and the comments based upon them are outside the scope of UKAS accreditation and are based on current SATRA knowledge.  
A satisfactory test report in no way implies that the product tested is approved by SATRA and no warranty is given as to the performance of the product tested. SATRA shall not be liable for any subsequent loss or damage incurred by the client as a result of information supplied in the report.  
The uncertainty of the results (UoM) in this report is based on a standard uncertainty multiplied by a coverage factor k=2, which provides a coverage probability of approximately 95%.

Report signed by: Lucy Cove  
Position: Technologist  
Department: Chemical & Analytical Technology  
(Page 1 of 10)

SATRA Technology Centre Ltd is a subsidiary of SATRA Ltd. Registered in England No. 3630206 at the above address.

**SATRA TECHNOLOGY**

**TECHNICAL REPORT**

**WORK REQUESTED:**  
Samples of gloves described as Disposable Powdered Latex Gloves referenced as MRJ01 were received on the 30<sup>th</sup> November 2020 for testing in accordance with EN 15523-1:2015+A1:2018 and assessment in accordance with the requirements of EN ISO 374-1:2016+A1:2018.

**SAMPLES SUBMITTED:**



Samples described as Disposable Powdered Latex Gloves referenced as MRJ01

**CONCLUSION:**  
When assessed in accordance with the requirements of EN ISO 374-1:2016+A1:2018 the samples of gloves described as Disposable Powdered Latex Gloves referenced as MRJ01 achieved the following performance levels:

Chemical	Performance level
40% Sodium hydroxide (CAS: 1310-73-2)	Inhomogeneous
30% Hydrogen peroxide (CAS: 7722-84-1)	2
37% Formaldehyde (CAS: 50-00-0)	Inhomogeneous

Full results are reported in the following tables.

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM030568/2048/LC/A  
Date: 12<sup>th</sup> January 2021 (Page 2 of 10)

Signed: Lucy Cove

**SATRA TECHNOLOGY**

**TECHNICAL REPORT**

**TESTING REQUIRED:**  
EN 15523-1:2015+A1:2018 - Determination of material resistance to permeation by chemicals - Part 1: Permeation by liquid chemical under conditions of continuous contact

**RESULTS AND REQUIREMENTS:**  
EN ISO 374-1:2016+A1:2018 - Protective gloves against dangerous chemicals and micro-organisms - Part 1: Terminology and performance requirements for chemical risks, Table 1: Permeation performance levels.

Permeation performance level	Measured breakthrough time (minutes)
1	>10
2	>30
3	>60
4	>120
5	>240
6	>480

Performance levels are based on the lowest individual result achieved per chemical.

Test specimens were washed in deionised water prior to conditioning in accordance with EN 15523-1:2015+A1:2018. This modification was to remove extractable substances from the gloves that otherwise might have interfered with the conductivity detection technique.

In accordance with clauses 8.5.1.2 and 8.5.1.4, the test results were outside the defined range and required an additional 3 specimens to be tested. All 6 results have been reported and the sample was found to be inhomogeneous.

In accordance with EN ISO 374-1 clause 5.4.1, the result is considered a fail due to non-homogeneity of the samples.

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM030568/2048/LC/A  
Date: 12<sup>th</sup> January 2021 (Page 3 of 10)

Signed: Lucy Cove

**SATRA TECHNOLOGY**

**TECHNICAL REPORT**

**Test/Property** Sample reference: Disposable Powdered Latex Gloves referenced as MRJ01

**Test information:** Chemical: 40% Sodium hydroxide  
Normalised permeation rate (NPR): 1 µg/cm<sup>2</sup>/min  
Conductivity  
Detection technique: (continuous measurement)  
Collection medium: Deionised water (closed loop)  
Collection medium stirring rate: 45 - 65 mm/min  
pH not constant to within ± 0.5%  
Test temperature: (23 ± 1) °C

**Performance:** Inhomogeneous

Specimen	Thickness (mm)	Breakthrough time (mins)
1	0.13	172
2	0.12	>480
3	0.12	>480
4	0.12	161
5	0.13	170
6	0.13	252

Test result: 161  
UoM: <1

Visual appearance of specimens after testing: Discoloured

Test specimens were washed in deionised water prior to conditioning in accordance with EN 15523-1:2015+A1:2018. This modification was to remove extractable substances from the gloves that otherwise might have interfered with the conductivity detection technique.

In accordance with clauses 8.5.1.2 and 8.5.1.4, the test results were outside the defined range and required an additional 3 specimens to be tested. All 6 results have been reported and the sample was found to be inhomogeneous.

In accordance with EN ISO 374-1 clause 5.4.1, the result is considered a fail due to non-homogeneity of the samples.


SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM030568/2048/LC/A  
Date: 12<sup>th</sup> January 2021 (Page 4 of 10)

Signed: Lucy Cove

**SATRA TECHNOLOGY**

**TECHNICAL REPORT**

**PERMEATION RATE GRAPHS**  
Challenge chemical: 40% Sodium Hydroxide



**PERMEATION RATE GRAPHS**  
Challenge chemical: 40% Sodium Hydroxide



SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM030568/2048/LC/A  
Date: 12<sup>th</sup> January 2021 (Page 5 of 10)

Signed: Lucy Cove

EN ISO21420 ISO374-3 乳胶



For SOP CAT-025, where both the  $P_1$  and  $P_n$  are observed in the same sampling range, uncertainty is expressed as the time difference between the mid-point of the range and the previous sampling time. This uncertainty is included in the reported result.



SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0305685/2049/LC/A  
Date: 12<sup>th</sup> January 2021 (Page 6 of 10)

Signed: *L. me*




For SOP CAT-025, where both the  $P_1$  and  $P_2$  are observed in the same sampling range, uncertainty is expressed as the time difference between the mid-point of the range and the previous sampling time.

In accordance with clauses 8.5.1.2 and 8.5.1.4, the test results were outside the defined range and required an additional 3 specimens to be tested. All 6 results have been reported and the sample was found to be inhomogeneous.

In accordance with EN ISO 374-1 clause 5.4.1, the result is considered a fail due to non-homogeneity of the samples.

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM305688/2049/LC/A  
Date: 13th January 2015 (Page 2 of 3)

Signed: *L. C. M.*



Formaldehyde is determined by discrete sampling; therefore the permeation rate graph is not a

o EN 16523-1:2015+A1:2016 does not require the test specimen thicknesses to be reported, this information is indicative only.

\* Breakthrough expressed as a range between discrete sampling points where the average permeation rate exceeds the NPR. Due to the complexity of the detection technique, the minimum sampling frequency as specified in table 1 of EN 16523-1:2015+A1:2016 is not possible.

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0305688/2049/LC/A  
Date: 12<sup>th</sup> January 2021 (Page 8 of 10)

Signed: *L. come*



## TERMS AND CONDITIONS FOR THE SALE OF GOODS AND/OR THE PROVISION OF SERVICES

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM03056882049/LC/A  
Date: 12<sup>th</sup> January 2021 (Page 9 of 11)

Signed: *L. come*

[illegible]

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0305688/2049LC/A  
Date: 12<sup>th</sup> January 2021 (Page 10 of 10)

Signed: *L. come*

EN ISO21420 ISO374-4 乳胶



Your reference:	CHT0305549
Date of report:	12 <sup>th</sup> January 2021
Samples received:	30 <sup>th</sup> November 2020
Date(s) work completed:	23 <sup>rd</sup> December 2020

**SATRA Technology Services (Dongguan) Ltd:**  
Customer: Fujian Jixiang Medical Technology Co., Ltd  
No. 29 Tingzhou Avenue,  
Cewu Town,  
Changling County,  
Fujian Province,  
China

**Conditions of Issue:**

This report may be forwarded to other parties provided that it is not changed in any way. It must not be published, for example by including it in advertisements, without the prior, written permission of SATRA.

Results given in this report refer only to the samples submitted for analysis and tested by SATRA. Comments are for guidance only.

Tests marked # fall outside the UKAS Accreditation Schedule for SATRA. All interpretations of results of such tests and the comments based upon them are outside the scope of UKAS accreditation and are based on current SATRA knowledge.

A satisfactory test report in no way implies that the product tested is approved by SATRA and no warranty is given as to the performance of the product tested. SATRA shall not be liable for any subsequent loss or damage incurred by the client as a result of information supplied in the report.

The uncertainty of the results (U<sub>95</sub>) in this report is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , which provides a coverage probability of approximately 95%.


Report signed by: Lucy Cove  
Position: Technologist  
Department: Chemical & Analytical Technology

SATRA Technology Centre Ltd is a subsidiary of SATRA. Registered in England No. 3856298 at the above address.



SAMPLE SUBMITTED

Samples of gloves described as Disposable Powdered Latex Gloves referenced as MRJ01 were received on the 30<sup>th</sup> November 2020 for testing in accordance with EN ISO 374-4:2019



When assessed in accordance with EN ISO 374-4:2019 the samples of gloves described as Disposable Powdered Latex Gloves referenced as MFJ01 achieved the following degradation results:

Chemical	Mean degradation / %
40% Sodium hydroxide (CAS: 1310-73-2)	-19.6
37% Formaldehyde (CAS: 50-00-0)	-18.6
30% Hydrogen peroxide (CAS: 7722-84-1)	21.3

- EN ISO 374-4:2019. Protective gloves against dangerous chemicals and micro-organisms. Part 4: Determination of resistance to degradation by chemicals.

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0305688/2049/LC/I  
Date: 12th January 2021

Figure 1



Sample description:	Disposable Powdered Latex Gloves referenced as MRJ31		
Challenge chemical:	40% Sodium hydroxide (CAS: 1310-73-2)		
Test temperature / °C:	(23 ± 1)		
Degradation / %:	Glove 1	Glove 2	Glove 3
	-18.0	-10.0	-30.7
Mean degradation (DR) / %:	-19.6		
Standard deviation ( $\sigma_m$ ) / %:	10.4		
UoM / % ±:	30.9		
Appearance of samples after testing:	Discoloured		

Sample description:	Disposable Powdered Latex Gloves referenced as MRJ2		
Challenge chemical:	37% Formaldehyde CAS: 50-00-0		
Test temperature / °C:	(23 ± 1)		
Degradation / %:	Glove 1	Glove 2	Glove 3
	-9.4	-16.8	-29.6
Mean degradation (DR) / %:	-18.6		
Standard deviation ( $\sigma_{\text{DR}}$ ) / %:	10.2		
UoM / ± %:	31.0		
Appearance of samples after testing:	Discoloured		

Sample description:	Disposable Powdered Latex Gloves referenced as MRJ01		
Challenge chemical:	30% Hydrogen peroxide (CAS: 7722-84-1)		
Test temperature / °C:	(23 ± 1)		
Degradation / %:	Glove 1	Glove 2	Glove 3
Mean degradation (DR) / %:	19.1	31.6	13.3
Standard deviation ( $\sigma_m$ ) / %:		21.3	9.3
UoM / %:		35.8	
Appearance of samples after testing:	Discoloured		

**NOTE:** Where the test specimens gave an increased puncture force after chemical exposure, the result is reported as a negative degradation.

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0305688/2049/LC1  
Date: 12<sup>th</sup> January 2021

Source:

[illegible]

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM0005688/2049/LC/B  
Date: 12th January 2021 (Page 4 of 4)

10

[illegible]

SATRA Technology Services (Dongguan) Ltd  
SATRA Reference: CHM030688/2049/LC/8  
Date: 12<sup>th</sup> January 2021

59

		 	
<h1 style="text-align: center;">Test Report</h1>			
Report No.: R019210225120	Date: May 05, 2021	Page 1 of 7	
<b>Applicant</b> 客户/申请人 Address : 福建泉州晋江闽华纺织有限公司 : No.29 Tangliang Avenue, Civea Town, Chong County, Fujian Province, China : 福建省晋江市陈埭镇溪寮村大厝前29号			
<b>Report on the submitted sample(s) told to be:</b> 提供检测样品及申报者所申报的样品情况如下。			
<b>Sample(s) Name</b> 样品名称 : NITRILE GLOVES : 丁腈手套 <b>Sample(s) resolved date</b> 样品接收日期 : Dec.20, 2020 <b>Product ID Number</b> : 2020-03-251 <b>Testing period</b> : From Dec.05, 2020 to Jan.06, 2021 <b>Test item/测试项</b> : 2020-01-251 to 2021-01-06.			
<b>Test Request</b> 客户要求		<b>Conclusion</b> 结论	
(a) As specified by client, to determine the Overall Migration in the submitted sample(s) in accordance with European Commission Regulation (EC) No 1935/2004 and European Commission Regulation (EU) 2017/2469 on rubber materials intended to come into contact with foodstuffs. 按照客户要求，参考欧盟委员会法规(EU) No.1935/2004 和欧洲委员会决议 (EU)2017/2469 与食品接触用橡胶材料。 是否欧盟委员会决议(EU)2017/2469		Pass 合格	
(b) As specified by client, to determine the Specific Migration of Primary Aromatic Amine in the submitted sample(s) in accordance with European Commission Regulation (EC) No 1935/2004 and European Commission Regulation (EU) 2017/2469 on rubber materials intended to come into contact with foodstuffs. 按照客户要求，参考欧盟委员会法规(EU) No.1935/2004 和欧洲委员会决议 (EU)2017/2469 与食品接触用橡胶材料。 是否欧盟委员会决议(EU)2017/2469		Pass 合格	
检测客户/申请人: 福建省晋江市闽华纺织有限公司 2020-03-251 和欧洲委员会决议 (EU)2017/2469 与食品接触用橡胶材料。 是否欧盟委员会决议(EU)2017/2469			
<div style="display: flex; justify-content: space-between; align-items: center;"> <div>  <div style="text-align: right;">   <b>Ben Ma</b>            Technical Manager            技术经理         </div> </div> </div>			
The Test Report is issued by the Company solely as a Certificate of Issuance of the Test Report prepared and/or submitted. The results shown on this Test Report apply only to the sample(s) submitted unless otherwise stated. This Test Report shall not be introduced as proof of fact, without written approval of the Company.			
本检测报告由本公司唯一作为检测证书发出，检测报告由本公司编制和/或提交。报告中所示的结果仅适用于提交的样品，除非另有说明。本检测报告不得作为事实的证明，未经本公司书面批准。			
<b>Shanghai General Standard Testing Service Co., Ltd</b> 上海经方检测技术有限公司 25A-3F No.146 Wuyang Road, Xujiaochang Road, Xujiaochang Street, Xujiaochang Town, Dongxihu, Shanghai, China 上海市嘉定区徐泾镇徐家昌路146号3层25A室			

**GST**

**BSI** **UKAS**  
BSI  
 BRITISH  
 STANDARDS  
 INSTITUTION  
UKAS  
 MANAGEMENT  
 SYSTEMS  
 CERTIFICATION

# Test Report

Report No.: RG87201225126

Date: Jun 07, 2012

Page 2 of 7

## Test Request

### 测试要求

## Conclusion

### 结论

Pass

(1) As specified by client, to determine the Specific Migration of Nitrosamine and Nitrosatable substances in the submitted sample(s) in accordance with European Commission Regulation (EC) No. 1935/2004 and Council of Europe Recommendation AT 2004(4) on safety materials intended to come into contact with foodstuffs.  
 检测客户产品，参考欧盟委员会法规(No. 1935/2004 和欧洲委员会决议(EUR/COM/04)与食品接触的橡胶材料，对样品进行迁移和迁移后添加物质迁移测试。

合格

## Photograph(s) of Sample

### 样品照片

GST authenticate the photo on original report only

此照片仅用于 GST 正本报告验证

## Specimen Description:

### 样品描述:

No.	Test party's name
01	测试样品名称
1	Purple gloves (Material: Butadiene)
	紫色手套 (材料: 丁腈)

This Test Report is issued by the Company solely in the Confidence of Issuance of the Report without prejudice or attached. The results shown in this Test Report only refer to the sample(s) tested unless otherwise stated. This Test Report should not be distributed outside or full without written approval of the Company.

Shenzhen General Standard Testing Service Co., Ltd

218A-1-5/F, Xing Wang Building, 18/19A-2nd industrial zone, Xixiang Street, Juehe Town, Shenzhen, Guangdong, China  
 Tel: 86-755-26011000 Website: www.gsbt.com

		 				
Report No.: JG073042201210		Date: Jan. 07, 2012				
标题: 結果:		Page 1 of 7				
(f) Overall Migration - European Commission Regulation (EC) No 1935/2004 and European Commission Recommendation 2015/2283 on rubber materials intended for use in contact with foodstuffs 食品接触用橡胶制品符合标准EN 1510:2007和欧洲委员会意见2015/2283关于食品接触的材料法规						
Test Methods: Web references: EN 1186-1:2007 & EN 1186-2:2007 & EN 1186-3:2002 & EN 1186-4:2002 测试方法: 参考文件: 标准EN 1186-1:2007和EN 1186-2:2007和EN 1186-3:2002和EN 1186-4:2002						
测试液	Unit	Time 时间	Temperature 温度	MDL 方法 检测限	Maximum Permissible Limit 最大允许限值	Results 结果
30%w/v acetic acid in aqueous solution 30%v/v 乙酸溶液	mg/dm <sup>2</sup>	24h	25°C	3	10	4.68
10%v/v ethanol in aqueous solution 10%v/v 乙醇溶液	mg/dm <sup>2</sup>	24h	25°C	3	10	5.34
20%v/v ethanol in aqueous solution 20%v/v 乙醇溶液	mg/dm <sup>2</sup>	24h	25°C	3	10	4.34
50%v/v ethanol in aqueous solution 50%v/v 乙醇溶液	mg/dm <sup>2</sup>	24h	25°C	3	10	5.53





## Test Report

Report No.: R03700022530

Date: 2021.06.01

Page: 4 of 7

(2) Specific Migration of Primary Aromatic Amine: European Commission Regulation (EC) No 1935/2004 and European Commission Recommendation R/2004/44 on rubber materials intended to come into contact with foodstuffs

特定芳香胺迁移试验: 欧盟委员会法规(CE) No 1935/2004和欧洲委员会决议 R/2004/44 与食品接触的橡胶材料

Test Method: With reference to BS EN 13136-1:2004 & DIN 53410:1995

测试方法: 参考标准: BS EN 13136-1:2004 & DIN 53410:1995

Test condition: 70% (w/v) acetic acid in aqueous solution, 25 °C, 2 hours

测试条件: 70% (w/v) 乙酸溶液, 25 °C, 2 小时

Test instrument: Gas Chromatography-Mass Spectrometer (GC-MS)

测试仪器: 气相色谱-质谱联用仪(GC-MS)

Test item(s) 测试项目	Unit 单位	MSX 方法依据	Maximum Permissible Limit 最大允许限量	Result(s) 结果
Specific Migration of Primary Aromatic Amine(PAA)	mg/kg	0.01	Not Detected 未检测到	1 N.D.
特定芳香胺迁移试验 (PAA)				

This Test Report is issued by the Company subject to its Conditions of Business of Test Report printed on/after or attached. The results shown in this Test Report (only) are the company's tested results otherwise stated. This Test Report shall be reproduced only in full, without omission of any of the Company.

2.8.18.15.07 Faw-Wing (Shanghai) Chemical Co., Ltd. Industrial Area, Xuying Street, Baotou, Shendong, Guangdong, China



# Test Report



Report No.: RGST20221210
Date: Jan. 07, 2021
Page: 7/7

(1) Specific Migration of Nitrosamines and N-nitrosatable Substances – European Commission Regulation (EC) No 1831/2003 and European Commission Regulation (EC) 2004/40 on rubber materials intended to come into contact with foodstuffs.

测试条件: 符合欧盟食品接触材料特定迁移 亚硝胺类物质迁移 (EC) No 1831/2003和欧洲委员会决议 (EC) 2004/40 食品接触材料迁移

Test Method: With reference to GB 19738-13 (3.34.1), 2008AHS EN 12668 2017

测试条件: 7% (w/v) acetic acid in aqueous solution, 40°C, 24 hours

测试条件: 7% (w/v) 乙酸水溶液, 40°C, 24 小时

Test Instrument: Liquid Chromatography-Mass Spectrometer (LC-MS)

测试仪器: 液相色谱-质谱联用仪

Test Item(s) 测试项目	Unit 单位	MEL 迁移限值 单位 mg/kg	Maximum Permissible Limit 最大允许量值	Results 结果
Migration of N-Nitrosamines: 亚硝胺类物质迁移	mg/kg	0.01	0.01	N.D.
Migration of N-Nitrosatable Substances 亚硝胺类物质迁移	mg/kg	0.1	0.1	N.D.

This Test Report is issued by the Company subject to the Conditions of Issuance of This Report printed/attached at attached.

The results shown in this Test Report only apply to the specific product unless otherwise stated. This Test Report does not represent opinion in itself, without written approval of the Company.

**Shanghai General Standard Testing Service Co., Ltd.**  
 上海通用标准测试服务有限公司  
 210007 Shanghai, China  
 210007 中国上海

# 欧盟委员会法规 (EC)1935/2004



<div><div><div><div><div><div><span></span></div><div><b>GST</b></div></div></div><div><div><div><span></span></div><div><span></span></div></div><div><div><span></span></div><div><span></span></div></div></div></div></div><div><div><div><div><div><span></span></div><div><b>GST</b></div></div></div><div><div><div><span></span></div><div><span></span></div></div></div><div><div><div><span></span></div><div><span></span></div></div></div></div></div></div> <div><div><div><div><div><span></span></div><div><b>GST</b></div></div></div><div><div><div><span></span></div><div><span></span></div></div></div><div><div><div><span></span></div><div><span></span></div></div></div></div></div>
--


# 我们的资质



福建集翔医疗科技有限公司

# 我们的资质

对外贸易经营者备案登记表			
备案登记编号: 02967774		统一社会信用代码: 91350821MA3HCY31G	
经营者中文名称: 福建集翔医疗科技有限公司		经营者英文名称: Fujian Jixiang Medical Technology Co., LTD	
组织机构代码	经营者类型: 有限责任公司		
住所	福建省长汀县武镇汀州大道南路29号		
经营场所(中文)	福建省长汀县武镇汀州大道南路29号		
经营场所(英文)	29# north tingzhou road, Cewu town, Changting County, Fujian province		
联系电话	18743008713	联系传真	
邮政编码	366300	电子邮箱	15039936879@163.com
工商登记注册日期	2020年04月14日	工商登记注册号	
依法办理工商登记的企业还应填写以下内容			
企业法定代表人姓名	张秀秀	有效证件号	35082219790126092X
注册资金	500万元		[万美元]
依法办理工商登记的外国(地区)企业或个体工商户(独资经营者)还应填写以下内容			
企业法定代表人姓名		有效证件号	
企业资产/个人财产			[万美元]
备注: 已完成原产地申请人备案, 请到检验检疫机构备案会办原产地证。			
填表前请认真阅读背面的条款, 并由企业法定代表人或个体工商户负责人签字、盖章。			
			
2020 年 09 月 14 日			

第二类医疗器械经营备案凭证	
企业名称	福建集翔医疗科技有限公司
法定代表人	张秀秀
企业负责人	张秀秀
经营方式	批发
住 所	福建省长汀县武镇汀州大道南路 29 号
经营场所	福建省长汀县武镇汀州大道南路 29 号
库房地址	福建省长汀县武镇汀州大道南路 29 号 1#厂房
经营范围	《医疗器械分类目录》(2002 版) II 类: 6820、6834、6864、6866、《医疗器械分类目录》(2017 版) II 类: 06、07、08、14、16、18。
备案部门(公章):  备案日期: 2020 年 09 月 14 日	

Fiscal Year 2021 CERTIFICATE OF FDA REGISTRATION			
This certifies that: <b>Fujian Jixiang Medical Technology Co., Ltd.</b> <b>29 Tingzhou Avenue South Road, Cewu Town, Changting County, Longyan, FUJIAN, 366399, CHINA</b> has completed the FDA Establishment Registration and Device Listing with the US Food & Drug Administration, through from a 3rd party registrant.			
Owner/Operator Number: 10078357			
Device Listing#:	Listing No.	Code:	Proprietary Name:
	D422378	QLB	Nitrile inspection gloves
			Latex examination gloves
			PVC inspection gloves
Activities: Manufacturer			
<p><small>will confirm that such registration remains effective upon request and presentation of this certificate until the end of the calendar year stated above, unless said registration is terminated after issuance of this certificate. 3rd party registrant makes no other representations or warranties, nor does this certificate make any representations or warranties to any person or entity other than the named certificate holder, for whose sole benefit it is issued. This certificate does not denote endorsement or approval of the certificate holder's device or establishment by the U.S. Food and Drug Administration. 3rd party registrant assumes no liability to any person or entity in connection with the foregoing.</small></p> <p><small>Pursuant to 21 CFR 807.26 "Registration of a device establishment or assignment of a registration number distinct in any way denote approval of the establishment or its products. Any representation that creates an impression of official approval because of registration or possession of a registration number is misleading and constitutes misbranding."</small></p> <p><small>The U.S. Food and Drug Administration does not issue a certificate of registration, nor does the U.S. Food and Drug Administration recognize a certificate of registration. U.S. 3rd party registrant is not affiliated with the U.S. Food and Drug Administration.</small></p>			
U.S. Agent Contact Name: Peimei Ding Address: 67 crescent Irvine, CA 92620 U.S.A. Phone: 1-949-2778689 E-mail: peimeiding@yahoo.com		 Issued Date: Oct 26, 2020 Expiration Date: December 31, 2021	

# 我们的资质

第一类医疗器械备案信息表	
备案号：闽龙械备 20200028 号	
备案人名称	福建集翔医疗科技有限公司
备案人组织机构代码	91350821MA34H3YN1G
备案人注册地址	福建省长汀县策武镇汀州大道南路 29 号
生产地址	福建省长汀县策武镇汀州大道南路 29 号
代理人	*****
代理人注册地址	*****
产品名称	医用手套检查手套
型号/规格	型号：无粉光面。 规格：S、M、L、XL。
产品描述	主要由丁腈橡胶原料加工制成，有足够的强度和阻隔性能，非无菌提供，一次性使用。
预期用途	用于戴在医生手上或者指上对患者病情进行检查或触诊。
备注	
备案单位 和日期	龙岩市市场监督管理局 备案日期：2020 年 12 月 24 日
变更情况	

第一类医疗器械备案凭证	
福建集翔医疗科技有限公司： 根据相关法规要求，对你单位第一类医疗器械：医用手 肘检查手套予以备案，备案号：闽龙械备 20200028 号。	
龙岩市市场监督管理局 备案日期：2020 年 12 月 24 日	

第一类医疗器械生产备案凭证				
备案号：闽龙食药监械生产备 202000008 号				
企业名称	福建集翔医疗科技有限公司			
住 所	福建省长汀县策武镇汀州大道南路 29 号			
生产场所	福建省长汀县策武镇汀州大道南路 29 号			
法定代表人	张秀秀	企业负责人	张秀秀	
生产范围	14-14 医护人员防护用品			
生产产品 列表	产品名称	产品备案号	是否 受托生产	备注
	14-14-04 医用手肘检查手套	闽龙械备 20200028 号	否	
龙岩市市场监督管理局 备案日期：2020 年 12 月 25 日				


**福建集翔**  
Fujian Jixiang


2020年01月04日  
(出具单位盖章)

**amfori**  **BSCI**  
Trade with purpose

This is an extract of the on line Audit Report. The complete report is available in the online BSCI Platform.  
Access [www.bscplatform.org](http://www.bscplatform.org) for entitled users only.

All rights reserved. No part of this publication may be reproduced, stored, or transmitted in any form or by any means, electronic, mechanical, photocopying, recording, or otherwise, without prior written permission from the publisher.





## EC Declaration of Conformity

Manufacturer:

Fujian Jixiang Medical Technology Co., Ltd.  
29 Ting Zhou Avenue South Road, Cewu Town,  
Chang Ting County, Fujian Province  
HongSong Xu  
Tel: +86-0597-6856569  
E-mail: 627913027@qq.com

whose single Authorized EU-Representative:

M/S CMC Medical Devices & Drugs S.L.  
C/Horacio Lengo N° 18 CP 29006, Málaga-Spain  
Tel: +34951214054  
Fax: +34952330100

We, the manufacturer, herewith declare that the products:

Disposable PVC Gloves  
PVC001

meet the provisions of Directive 93/42/EEC which apply to them.

The medical device has been assigned to class I according to Annex IX of the Directive 93/42/EEC. It bears the mark



following the procedure relating to the EC Declaration of Conformity set out in Annex VII of Directive 93/42/EEC.

This Declaration of Conformity covers all medical devices as specified in the product list belonging to this declaration.

The above mentioned declaration of conformity is exclusively under the responsibility of

Fujian Jixiang Medical Technology Co., Ltd.  
29 Ting Zhou Avenue South Road, Cewu Town, Chang Ting County, Fujian Province

2020.11.19

Place, date

Legally binding signature, Function

## EC Declaration of Conformity

Manufacturer:

Fujian Jixiang Medical Technology Co., Ltd.  
29 Ting Zhou Avenue South Road, Cewu Town,  
Chang Ting County, Fujian Province  
HongSong Xu  
Tel: +86-0597-6856569  
E-mail: 627913027@qq.com

whose single Authorized EU-Representative:

M/S CMC Medical Devices & Drugs S.L.  
C/Horacio Lengo N° 18 CP 29006, Málaga-Spain  
Tel: +34951214054  
Fax: +34952330100

We, the manufacturer, herewith declare that the products:

Latex Gloves  
RL001

meet the provisions of Directive 93/42/EEC which apply to them.

The medical device has been assigned to class I according to Annex IX of the Directive 93/42/EEC. It bears the mark



following the procedure relating to the EC Declaration of Conformity set out in Annex VII of Directive 93/42/EEC.

This Declaration of Conformity covers all medical devices as specified in the product list belonging to this declaration.

The above mentioned declaration of conformity is exclusively under the responsibility of

Fujian Jixiang Medical Technology Co., Ltd.  
29 Ting Zhou Avenue South Road, Cewu Town, Chang Ting County, Fujian Province

2020.11.19

Place, date

Legally binding signature, Function

## EC Declaration of Conformity

Manufacturer:

Fujian Jixiang Medical Technology Co., Ltd.  
29 Ting Zhou Avenue South Road, Cewu Town,  
Chang Ting County, Fujian Province  
HongSong Xu  
Tel: +86-0597-6856569  
E-mail: 627913027@qq.com

whose single Authorized EU-Representative:

M/S CMC Medical Devices & Drugs S.L.  
C/Horacio Lengo N° 18 CP 29006, Málaga-Spain  
Tel: +34951214054  
Fax: +34952330100

We, the manufacturer, herewith declare that the products:

Nitrile Gloves  
DJ001

meet the provisions of Directive 93/42/EEC which apply to them.

The medical device has been assigned to class I according to Annex IX of the Directive 93/42/EEC. It bears the mark



following the procedure relating to the EC Declaration of Conformity set out in Annex VII of Directive 93/42/EEC.

This Declaration of Conformity covers all medical devices as specified in the product list belonging to this declaration.

The above mentioned declaration of conformity is exclusively under the responsibility of

Fujian Jixiang Medical Technology Co., Ltd.  
29 Ting Zhou Avenue South Road, Cewu Town, Chang Ting County, Fujian Province

2020.11.19

Place, date

Legally binding signature, Function

## EC REP CERTIFICATE



CMC MEDICAL DEVICES & DRUGS SL  
NO. CMC/CE/2020/30112020.10

CONFIRMED THAT CMC MEDICAL DEVICES & DRUGS S.L. is the European Authorized Representative of

**Fujian Jixiang Medical Technology Co., Ltd**  
29 Ting Zhou Avenue South Road, Cewu Town,  
Chang Ting County, Fujian Province

The certificate remains valid until the expiration agreement of EC REP, manufacturing conditions, the quality system or relevant legislation are changed. The validity is conditioned by positive results of periodic surveillance audits.

The product liability rests with the manufacturer in accordance with applicable directive and standard, after fulfilling of the relevant EU legislation requirements, the manufacturer shall affix relevant CE marking to all above mentioned models of the medical device.

Complies with the applicable essential requirements of the council directive 93/42/EEC on medical devices as amended.

The products in Annex I was registered in Spanish MOH with number RPS/2755/2020



Issued on: 30/11/2020



Authorized Signatory  
CMC Medical Devices & Drugs SL

Valid until: 29/11/2021

[www.cmcmedicaldevices.com](http://www.cmcmedicaldevices.com)

## EC REP CERTIFICATE



ANNEX I Medical Device Product Description

Nitrile Gloves  
Model: DJ001

Latex Gloves  
Model: RJ001

Disposable PVC Gloves  
Model: PVC001



[www.cmcmedicaldevices.com](http://www.cmcmedicaldevices.com)



**THANKS!!!**